


7471A

LB 86764
INSTRUMENT ID: CV2

Linear

 μ Abs.:

42824



A= 0.0000e+000

B= 2.3564e-004

C= 1.2236e-002

Rho= 0.9997837

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	19	0.000	19				
0.2	0.200	0.209	0.009	837	0.0 %	837				
2.5	2.500	2.515	0.015	10620	0.0 %	10620				
5.0	5.000	5.010	0.010	21210	0.0 %	21210				
7.5	7.500	7.345	-0.155	31119	0.0 %	31119				
10.0	10.000	10.103	0.103	42824	0.0 %	42824				

LB86764 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 07 Apr 2017 14:11:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	19	-	0.0000	7471A	PPB	07 Apr 2017 14:22:43	S	Std
0.2 - 1	50.2	837	-	0.2000	7471A	PPB	07 Apr 2017 14:25:00	S	Std
2.5 - 1	52.5	10620	-	2.5000	7471A	PPB	07 Apr 2017 14:27:18	S	Std
5.0 - 1	55	21210	-	5.0000	7471A	PPB	07 Apr 2017 14:29:36	S	Std
7.5 - 1	57.5	31119	-	7.5000	7471A	PPB	07 Apr 2017 14:31:53	S	Std
10.0 - 1	510	42824	-	10.0000	7471A	PPB	07 Apr 2017 14:34:09	S	Std
ICV11 - 1	ICV11	16510	3.9027	-	7471A	PPB	07 Apr 2017 14:36:59	U	SMPL
ICB11 - 1	ICB11	-67	-0.0036	-	7471A	PPB	07 Apr 2017 14:39:14	U	SMPL
CCV80 - 1	CCV80	21164	4.9994	-	7471A	PPB	07 Apr 2017 14:43:20	U	SMPL
CCB80 - 1	CCB80	-54	-0.0005	-	7471A	PPB	07 Apr 2017 14:45:36	U	SMPL
CRA - 1	CRA	841	0.2104	-	7471A	PPB	07 Apr 2017 14:47:51	U	SMPL
HighStd - 1	HighStd	43189	10.1895	-	7471A	PPB	07 Apr 2017 14:50:07	U	SMPL
ChkStd - 1	ChkStd	29330	6.9237	-	7471A	PPB	07 Apr 2017 14:52:23	U	SMPL
PB98115BL - 1	PBS	-112	-0.0142	-	7471A	PPB	07 Apr 2017 14:54:40	U	SMPL
PB98115BS - 1	LCSS	14512	3.4319	-	7471A	PPB	07 Apr 2017 14:56:56	U	SMPL
I2574-01 - 1	B2-S1(0-2)	4022	0.9600	-	7471A	PPB	07 Apr 2017 14:59:13	U	SMPL
I2574-02 - 1	B2-S2(2-4)	935	0.2326	-	7471A	PPB	07 Apr 2017 15:01:31	U	SMPL
I2574-03 - 1	B2-S3(4-6)	860	0.2149	-	7471A	PPB	07 Apr 2017 15:03:48	U	SMPL
I2574-04 - 1	B19-S1(2-4)	1400	0.3421	-	7471A	PPB	07 Apr 2017 15:06:06	U	SMPL
I2574-05 - 1	B57-S1(0-2)	25822	6.0971	-	7471A	PPB	07 Apr 2017 15:08:22	U	SMPL
CCV81 - 1	CCV81	21283	5.0275	-	7471A	PPB	07 Apr 2017 15:10:38	U	SMPL
CCB81 - 1	CCB81	-28	0.0056	-	7471A	PPB	07 Apr 2017 15:12:54	U	SMPL
I2574-06 - 1	B45-S1(0-2)	947	0.2354	-	7471A	PPB	07 Apr 2017 15:15:10	U	SMPL
I2574-07 - 1	B49-S1(0-2)	970	0.2408	-	7471A	PPB	07 Apr 2017 15:17:26	U	SMPL
I2574-08 - 1	B50-S1(0-2)	6420	1.5251	-	7471A	PPB	07 Apr 2017 15:19:42	U	SMPL
I2574-09 - 1	B44-S1(0-2)	2234	0.5387	-	7471A	PPB	07 Apr 2017 15:21:58	U	SMPL
I2574-09DUP - 1	B44-S1(0-2)DUP	2344	0.5646	-	7471A	PPB	07 Apr 2017 15:24:15	U	SMPL
I2574-09MS - 1	B44-S1(0-2)MS	17297	4.0882	-	7471A	PPB	07 Apr 2017 15:26:32	U	SMPL
I2574-09MSD - 1	B44-S1(0-2)MSD	17046	4.0290	-	7471A	PPB	07 Apr 2017 15:28:49	U	SMPL
I2574-09LX5 - 1		408	0.1084	-	7471A	PPB	07 Apr 2017 15:31:07	U	SMPL
I2574-09A - 1		20091	4.7466	-	7471A	PPB	07 Apr 2017 15:33:25	U	SMPL
CCV82 - 1	CCV82	21334	5.0395	-	7471A	PPB	07 Apr 2017 15:35:41	U	SMPL
CCB82 - 1	CCB82	-22	0.0071	-	7471A	PPB	07 Apr 2017 15:37:58	U	SMPL