


245.1

LB 86772
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35811



A= 0.0000e+000

B= 2.8144e-004

C= 1.4586e-002

Rho= 0.9998319

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	14	0.000	14				
0.2	0.200	0.213	0.013	705	0.0 %	705				
2.5	2.500	2.501	0.001	8833	0.0 %	8833				
5.0	5.000	5.010	0.010	17748	0.0 %	17748				
7.5	7.500	7.365	-0.135	26118	0.0 %	26118				
10.0	10.000	10.093	0.093	35811	0.0 %	35811				

LB86772 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 Apr 2017 15:46:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	14	-	0.0000	245.1	PPB	07 Apr 2017 15:47:12	S	Std
0.2 - 1	50.2	705	-	0.2000	245.1	PPB	07 Apr 2017 15:49:20	S	Std
2.5 - 1	52.5	8833	-	2.5000	245.1	PPB	07 Apr 2017 15:51:28	S	Std
5.0 - 1	55	17748	-	5.0000	245.1	PPB	07 Apr 2017 15:53:37	S	Std
7.5 - 1	57.5	26118	-	7.5000	245.1	PPB	07 Apr 2017 15:55:45	S	Std
10.0 - 1	510	35811	-	10.0000	245.1	PPB	07 Apr 2017 15:57:52	S	Std
ICV12 - 1	ICV12	14137	3.9933	-245.1		PPB	07 Apr 2017 16:00:33	U	SMPL
ICB12 - 1	ICB12	-24	0.0078	-245.1		PPB	07 Apr 2017 16:02:39	U	SMPL
CCV83 - 1	CCV83	17806	5.0259	-245.1		PPB	07 Apr 2017 16:04:45	U	SMPL
CCB83 - 1	CCB83	-21	0.0087	-245.1		PPB	07 Apr 2017 16:06:52	U	SMPL
CRA - 1	CRA	697	0.2107	-245.1		PPB	07 Apr 2017 16:08:59	U	SMPL
HighStd - 1	HighStd	35513	10.0093	-245.1		PPB	07 Apr 2017 16:11:05	U	SMPL
ChkStd - 1	ChkStd	24454	6.8969	-245.1		PPB	07 Apr 2017 16:13:12	U	SMPL
PB98139BL - 1	PBW	-51	0.0002	-245.1		PPB	07 Apr 2017 16:15:20	U	SMPL
PB98139BS - 1	LCSW	13686	3.8664	-245.1		PPB	07 Apr 2017 16:20:02	U	SMPL
I2573-01 - 1	COMP	454	0.1424	-245.1		PPB	07 Apr 2017 16:22:09	U	SMPL
I2573-01DUP - 1	COMPDUP	390	0.1243	-245.1		PPB	07 Apr 2017 16:24:17	U	SMPL
I2573-01MS - 1	COMPMS	13752	3.8849	-245.1		PPB	07 Apr 2017 16:28:48	U	SMPL
I2573-01MSD - 1	COMPMSD	13810	3.9013	-245.1		PPB	07 Apr 2017 16:32:31	U	SMPL
I2589-01 - 1	040517-DSDC-E-D-M	-24	0.0078	-245.1		PPB	07 Apr 2017 16:34:37	U	SMPL
CCV84 - 1	CCV84	17961	5.0695	-245.1		PPB	07 Apr 2017 16:36:44	U	SMPL
CCB84 - 1	CCB84	-29	0.0064	-245.1		PPB	07 Apr 2017 16:38:51	U	SMPL
I2589-02 - 1	040517-DSDC-E-U-M	22	0.0208	-245.1		PPB	07 Apr 2017 16:40:58	U	SMPL
I2589-03 - 1	040517-PRRT-E-D-M	18	0.0197	-245.1		PPB	07 Apr 2017 16:43:05	U	SMPL
I2590-01 - 1	040517-PRRT-E-U-M	32	0.0236	-245.1		PPB	07 Apr 2017 16:45:11	U	SMPL
I2573-01LX5 - 1		105	0.0441	-245.1		PPB	07 Apr 2017 16:47:18	U	SMPL
I2573-01A - 1		14444	4.0797	-245.1		PPB	07 Apr 2017 16:49:26	U	SMPL
CCV85 - 1	CCV85	17694	4.9944	-245.1		PPB	07 Apr 2017 16:51:34	U	SMPL
CCB85 - 1	CCB85	-36	0.0045	-245.1		PPB	07 Apr 2017 16:53:42	U	SMPL