



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB86800

Temp.: 23.4 °C

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/10/2017	17:30
2	CAL2	1	Water			7.00	04/10/2017	17:31
3	CAL3	1	Water			10.03	04/10/2017	17:32
4	ICV	1	Water			7.01	04/10/2017	17:33
5	CCV1	1	Water			2.01	04/10/2017	17:34
6	I2619-02	1	Water			7.51	04/11/2017	17:35
7	I2619-02DUP	1	Water			7.52	04/10/2017	17:36
8	I2621-02	1	Water			7.38	04/10/2017	17:37
9	CCV2	1	Water			12.02	04/10/2017	17:39

2386800

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I2621

WorkList ID : 97297

Date : 4/10/2017 5:28:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/10/2017	Water	I2619-02	Corrosivity	Cool 4 deg C	PSEG01	O41	SUMP	04/10/2017	9040C
	Water	I2621-02	Corrosivity	Cool 4 deg C	SEAC01	O23	GROUND-WATER	04/10/2017	9040C

Date/Time 04-10-17 5:28 PM
Received by: JP
Relinquished by: GP

Date/Time 04-10-17 6:35 PM
Received by: GP
Relinquished by: JP