



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86801

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/10/2017	17:30
2	CAL2	1	Water			7.00	04/10/2017	17:31
3	CAL3	1	Water			10.03	04/10/2017	17:32
4	ICV	1	Water			7.01	04/10/2017	17:33
5	CCV1	1	Water			2.02	04/10/2017	17:45
6	I2620-05	1	Solid	20.02	20	8.03	04/10/2017	17:46
7	I2620-05DUP	1	Solid	20.04	20	8.04	04/10/2017	17:47
8	I2620-10	1	Solid	20.05	20	6.95	04/10/2017	17:48
9	I2620-15	1	Solid	20.03	20	6.49	04/10/2017	17:50
10	I2621-01	1	Solid	20.01	20	8.75	04/10/2017	17:51
11	CCV2	1	Water			12.02	04/10/2017	17:52

2386801

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITYS I2621

WorkList ID : 97298

Date : 4/10/2017 5:29:15 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/13/2017	Solid	I2620-05	Corrosivity	Cool 4 deg C	PSEG01		D001-D002-D003	04/10/2017	9045C
04/13/2017	Solid	I2620-10	Corrosivity	Cool 4 deg C	PSEG01		D004-D005-D006	04/10/2017	9045C
04/13/2017	Solid	I2620-15	Corrosivity	Cool 4 deg C	PSEG01		SOUTH-MAHWAH	04/10/2017	9045C
	Solid	I2621-01	Corrosivity	Cool 4 deg C	SEAC01	O23	SOIL-CUTTINGS	04/10/2017	9045C

Date/Time 04-10-17 5:28 PM
Received by: JP
Relinquished by: CP

Date/Time 04-10-17 6:35 PM
Received by: CP
Relinquished by: JP