

7470A


L1386807

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33822



A= 0.0000e+000

B= 2.9823e-004

C= -2.1146e-002

Rho= 0.9998345

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	50	0.000	50				
0.2	0.200	0.195	-0.005	724	0.0 %	724				
2.5	2.500	2.601	0.101	8793	0.0 %	8793				
5.0	5.000	4.908	-0.092	16528	0.0 %	16528				
7.5	7.500	7.437	-0.063	25008	0.0 %	25008				
10.0	10.000	10.065	0.065	33822	0.0 %	33822				

LB86807 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Apr 2017 09:59:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	50	-	0.0000	7470A	PPB	11 Apr 2017 10:23:52	S	Std
0.2 - 1	50.2	724	-	0.2000	7470A	PPB	11 Apr 2017 10:26:00	S	Std
2.5 - 1	52.5	8793	-	2.5000	7470A	PPB	11 Apr 2017 10:28:09	S	Std
5.0 - 1	55	16528	-	5.0000	7470A	PPB	11 Apr 2017 10:30:17	S	Std
7.5 - 1	57.5	25008	-	7.5000	7470A	PPB	11 Apr 2017 10:32:24	S	Std
10.0 - 1	510	33822	-	10.0000	7470A	PPB	11 Apr 2017 10:34:31	S	Std
ICV13 - 1	ICV13	13028	3.8641	-	7470A	PPB	11 Apr 2017 10:37:08	U	SMPL
ICB13 - 1	ICB13	-72	-0.0426	-	7470A	PPB	11 Apr 2017 10:39:14	U	SMPL
CCV86 - 1	CCV86	16567	4.9196	-	7470A	PPB	11 Apr 2017 10:41:20	U	SMPL
CCB86 - 1	CCB86	-81	-0.0453	-	7470A	PPB	11 Apr 2017 10:43:27	U	SMPL
CRA - 1	CRA	700	0.1876	-	7470A	PPB	11 Apr 2017 10:45:34	U	SMPL
HighStd - 1	HighStd	33893	10.0866	-	7470A	PPB	11 Apr 2017 10:47:40	U	SMPL
ChkStd - 1	ChkStd	22923	6.8151	-	7470A	PPB	11 Apr 2017 10:49:47	U	SMPL
PB98165BL - 1	PBW	-130	-0.0599	-	7470A	PPB	11 Apr 2017 10:51:55	U	SMPL
PB98165BS - 1	LCSW	11997	3.5567	-	7470A	PPB	11 Apr 2017 10:54:04	U	SMPL
PB98165TB - 1	PB98140TB	-69	-0.0417	-	7470A	PPB	11 Apr 2017 10:56:12	U	SMPL
I2582-02 - 1	GW-1	1	-0.0208	-	7470A	PPB	11 Apr 2017 10:58:20	U	SMPL
I2582-05 - 1	GW-2	22	-0.0146	-	7470A	PPB	11 Apr 2017 11:00:28	U	SMPL
I2587-04 - 1	INKWASH	163	0.0275	-	7470A	PPB	11 Apr 2017 11:02:36	U	SMPL
I2595-03 - 1	GRID-1	0	-0.0211	-	7470A	PPB	11 Apr 2017 11:04:43	U	SMPL
CCV87 - 1	CCV87	16550	4.9145	-	7470A	PPB	11 Apr 2017 11:06:50	U	SMPL
CCB87 - 1	CCB87	-37	-0.0322	-	7470A	PPB	11 Apr 2017 11:08:56	U	SMPL
I2595-06 - 1	GRID-4	4	-0.0200	-	7470A	PPB	11 Apr 2017 11:11:03	U	SMPL
I2595-08 - 1	A2394	16	-0.0164	-	7470A	PPB	11 Apr 2017 11:13:09	U	SMPL
I2595-08DUP - 1	A2394DUP	15	-0.0167	-	7470A	PPB	11 Apr 2017 11:15:16	U	SMPL
I2595-08MS - 1	A2394MS	13621	4.0410	-	7470A	PPB	11 Apr 2017 11:17:23	U	SMPL
I2595-08MSD - 1	A2394MSD	13749	4.0792	-	7470A	PPB	11 Apr 2017 11:19:31	U	SMPL
I2595-08LX5 - 1		-62	-0.0396	-	7470A	PPB	11 Apr 2017 11:21:38	U	SMPL
I2595-08A - 1		14788	4.3890	-	7470A	PPB	11 Apr 2017 11:23:46	U	SMPL
CCV88 - 1	CCV88	16622	4.9360	-	7470A	PPB	11 Apr 2017 11:25:55	U	SMPL
CCB88 - 1	CCB88	-86	-0.0468	-	7470A	PPB	11 Apr 2017 11:28:03	U	SMPL