


7471A

LB86811
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45069



A= 0.0000e+000

B= 2.2401e-004

C= 1.2404e-002

Rho= 0.9998058

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	77	0.000	77				
0.2	0.200	0.222	0.022	935	0.0 %	935				
2.5	2.500	2.492	-0.008	11070	0.0 %	11070				
5.0	5.000	4.984	-0.016	22193	0.0 %	22193				
7.5	7.500	7.364	-0.136	32821	0.0 %	32821				
10.0	10.000	10.108	0.108	45069	0.0 %	45069				

LB86811 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 11 Apr 2017 11:42:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.2	77	-	0.00007471A	PPB		11 Apr 2017 11:46:04	S	Std
0.2 - 1	52.5	935	-	0.20007471A	PPB		11 Apr 2017 11:48:21	S	Std
2.5 - 1	55	11070	-	2.50007471A	PPB		11 Apr 2017 11:50:39	S	Std
5.0 - 1	57.5	22193	-	5.00007471A	PPB		11 Apr 2017 11:52:56	S	Std
7.5 - 1	510	32821	-	7.50007471A	PPB		11 Apr 2017 11:55:14	S	Std
10.0 - 1		45069	-	10.00007471A	PPB		11 Apr 2017 11:57:30	S	Std
ICV14 - 1	ICV14	17374	3.9043	-7471A	PPB		11 Apr 2017 12:00:20	U	SMPL
ICB14 - 1	ICB14	-6	0.0111	-7471A	PPB		11 Apr 2017 12:02:35	U	SMPL
CCV89 - 1	CCV89	21711	4.8758	-7471A	PPB		11 Apr 2017 12:04:51	U	SMPL
CCB89 - 1	CCB89	-3	0.0117	-7471A	PPB		11 Apr 2017 12:07:07	U	SMPL
CRA - 1	CRA	944	0.2239	-7471A	PPB		11 Apr 2017 12:09:23	U	SMPL
HighStd - 1	HighStd	45002	10.0931	-7471A	PPB		11 Apr 2017 12:11:38	U	SMPL
ChkStd - 1	ChkStd	30543	6.8542	-7471A	PPB		11 Apr 2017 12:13:55	U	SMPL
PB98172BL - 1	PBS	-2	0.0120	-7471A	PPB		11 Apr 2017 12:16:12	U	SMPL
PB98172BS - 1	LCSS	15103	3.3956	-7471A	PPB		11 Apr 2017 12:18:28	U	SMPL
I2595-01 - 1	GRID-1	4302	0.9761	-7471A	PPB		11 Apr 2017 12:20:45	U	SMPL
I2595-04 - 1	GRID-4	3139	0.7156	-7471A	PPB		11 Apr 2017 12:23:02	U	SMPL
I2598-01 - 1	B-47-S1(2-4)	1909	0.4400	-7471A	PPB		11 Apr 2017 12:25:20	U	SMPL
I2598-01DUP - 1	B-47-S1(2-4)DUP	1609	0.3728	-7471A	PPB		11 Apr 2017 12:27:38	U	SMPL
I2598-01MS - 1	B-47-S1(2-4)MS	16887	3.7952	-7471A	PPB		11 Apr 2017 12:29:54	U	SMPL
CCV90 - 1	CCV90	21759	4.8865	-7471A	PPB		11 Apr 2017 12:32:10	U	SMPL
CCB90 - 1	CCB90	25	0.0180	-7471A	PPB		11 Apr 2017 12:34:26	U	SMPL
I2598-01MSD - 1	B-47-S1(2-4)MSD	17162	3.8568	-7471A	PPB		11 Apr 2017 12:36:42	U	SMPL
I2598-02 - 1	B-61-S2(8-10)	1202	0.2817	-7471A	PPB		11 Apr 2017 12:38:58	U	SMPL
I2598-03 - 1	B-56-S1(0-2)	1530	0.3551	-7471A	PPB		11 Apr 2017 12:41:13	U	SMPL
I2598-04 - 1	B-64-S1(0-2)	5041	1.1416	-7471A	PPB		11 Apr 2017 12:43:29	U	SMPL
I2598-05 - 1	B-64-S2(8-10)	1365	0.3182	-7471A	PPB		11 Apr 2017 12:45:46	U	SMPL
I2598-06 - 1	B-62-S1(0-2)	2010	0.4627	-7471A	PPB		11 Apr 2017 12:48:03	U	SMPL
I2598-07 - 1	B-62-S2(8-10)	1973	0.4544	-7471A	PPB		11 Apr 2017 12:50:20	U	SMPL
I2598-08 - 1	B-25-S1(0-2)	1698	0.3928	-7471A	PPB		11 Apr 2017 12:52:38	U	SMPL
I2598-09 - 1	B-25-S2(13-15)	933	0.2214	-7471A	PPB		11 Apr 2017 12:54:56	U	SMPL
I2598-10 - 1	B-63-S1(0-2)	1036	0.2445	-7471A	PPB		11 Apr 2017 12:57:12	U	SMPL
CCV91 - 1	CCV91	22101	4.9631	-7471A	PPB		11 Apr 2017 12:59:28	U	SMPL
CCB91 - 1	CCB91	25	0.0180	-7471A	PPB		11 Apr 2017 13:01:45	U	SMPL
I2598-11 - 1	B-63-S2(8-10)	6541	1.4776	-7471A	PPB		11 Apr 2017 13:04:01	U	SMPL
I2610-01 - 1	SOURCE3-SE	8047	1.8150	-7471A	PPB		11 Apr 2017 13:06:18	U	SMPL
PB98173BL - 1	PBS	87	0.0319	-7471A	PPB		11 Apr 2017 13:08:34	U	SMPL
PB98173BS - 1	LCSS	14901	3.3503	-7471A	PPB		11 Apr 2017 13:10:51	U	SMPL
I2601-01 - 1	B-31-S1(6-8)	4237	0.9615	-7471A	PPB		11 Apr 2017 13:13:07	U	SMPL
I2601-02 - 1	B-31-S2(14-16)	186	0.0541	-7471A	PPB		11 Apr 2017 13:15:24	U	SMPL
I2601-03 - 1	B-35-S1(0-2)	3552	0.8081	-7471A	PPB		11 Apr 2017 13:17:41	U	SMPL
I2601-04 - 1	B-30-S1(0-2)	1143	0.2684	-7471A	PPB		11 Apr 2017 13:19:59	U	SMPL
I2601-05 - 1	B-30-S2(10-12)	1490	0.3462	-7471A	PPB		11 Apr 2017 13:22:17	U	SMPL
I2601-06 - 1	B-32-S1(0-2)	6212	1.4039	-7471A	PPB		11 Apr 2017 13:24:33	U	SMPL
CCV92 - 1	CCV92	22241	4.9945	-7471A	PPB		11 Apr 2017 13:26:49	U	SMPL
CCB92 - 1	CCB92	20	0.0169	-7471A	PPB		11 Apr 2017 13:29:06	U	SMPL
I2601-07 - 1	B-28-S1(4-6)	1629	0.3773	-7471A	PPB		11 Apr 2017 13:31:22	U	SMPL
I2601-08 - 1	B-28-S2(10-12)	702	0.1697	-7471A	PPB		11 Apr 2017 13:33:39	U	SMPL
I2601-09 - 1	B-26-S1(0-2)	1076	0.2534	-7471A	PPB		11 Apr 2017 13:35:55	U	SMPL
I2601-10 - 1	B-26-S2(12-12.4)	324	0.0850	-7471A	PPB		11 Apr 2017 13:38:11	U	SMPL
I2601-11 - 1	B-27-S1(2-4)	1662	0.3847	-7471A	PPB		11 Apr 2017 13:40:28	U	SMPL
I2601-12 - 1	B-27-S2(14-16)	334	0.0872	-7471A	PPB		11 Apr 2017 13:42:45	U	SMPL
I2601-13 - 1	B-29-S1(0-2)	2671	0.6107	-7471A	PPB		11 Apr 2017 13:45:03	U	SMPL
I2601-14 - 1	B-29-S2(10-12)	892	0.2122	-7471A	PPB		11 Apr 2017 13:47:20	U	SMPL
I2601-15 - 1	B-60-S1(2-4)	2180	0.5007	-7471A	PPB		11 Apr 2017 13:49:38	U	SMPL
I2601-16 - 1	B-60-S2(6-8)	1639	0.3795	-7471A	PPB		11 Apr 2017 13:51:54	U	SMPL
CCV93 - 1	CCV93	22413	5.0330	-7471A	PPB		11 Apr 2017 13:54:11	U	SMPL
CCB93 - 1	CCB93	12	0.0151	-7471A	PPB		11 Apr 2017 13:56:27	U	SMPL
I2601-17 - 1	B-58-S1(0-2)	889	0.2115	-7471A	PPB		11 Apr 2017 13:58:44	U	SMPL
I2601-18 - 1	B-59-S1(0-2)	2692	0.6154	-7471A	PPB		11 Apr 2017 14:01:00	U	SMPL
I2601-19 - 1	B-59-S2(8-10)	1750	0.4044	-7471A	PPB		11 Apr 2017 14:03:16	U	SMPL
I2601-19DUP - 1	B-59-S2(8-10)DUP	1671	0.3867	-7471A	PPB		11 Apr 2017 14:05:33	U	SMPL
I2601-19MS - 1	B-59-S2(8-10)MS	16982	3.8165	-7471A	PPB		11 Apr 2017 14:07:50	U	SMPL
I2601-19MSD - 1	B-59-S2(8-10)MSD	16909	3.8001	-7471A	PPB		11 Apr 2017 14:10:06	U	SMPL
I2601-20 - 1	B-61-S1(0-2)	4890	1.1078	-7471A	PPB		11 Apr 2017 14:12:24	U	SMPL
CCV94 - 1	CCV94	22386	5.0270	-7471A	PPB		11 Apr 2017 14:14:41	U	SMPL
CCB94 - 1	CCB94	22	0.0173	-7471A	PPB		11 Apr 2017 14:16:59	U	SMPL
I2598-01LX5 - 1		427	0.1081	-7471A	PPB		11 Apr 2017 14:19:15	U	SMPL
I2598-01A - 1		19236	4.3214	-7471A	PPB		11 Apr 2017 14:21:32	U	SMPL
I2601-19LX5 - 1		379	0.0973	-7471A	PPB		11 Apr 2017 14:23:48	U	SMPL
I2601-19A - 1		19305	4.3368	-7471A	PPB		11 Apr 2017 14:26:05	U	SMPL
CCV95 - 1	CCV95	22417	5.0339	-7471A	PPB		11 Apr 2017 14:28:21	U	SMPL
CCB95 - 1	CCB95	28	0.0187	-7471A	PPB		11 Apr 2017 14:30:38	U	SMPL