



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86813

Temp.: 21.2 °C

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/11/2017	11:50
2	CAL2	1	Water			7.00	04/11/2017	11:51
3	CAL3	1	Water			10.04	04/11/2017	11:52
4	ICV	1	Water			7.01	04/11/2017	11:53
5	CCV1	1	Water			2.02	04/11/2017	11:54
6	I2628-01	1	Water			4.69	04/11/2017	11:55
7	I2628-01DUP	1	Water			4.70	04/11/2017	11:56
8	CCV2	1	Water			12.02	04/11/2017	11:57

WORKLIST(Hardcopy Internal Chain)

LB 86813

WorkList Name : ph i2628 WorkList ID : 97336 Date : 4/11/2017 11:43:18 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2628-01	pH	Cool 4 deg C	STER11	O13	POOLWATER	04/05/2017	SM 4500-PH B

Date/Time 04-11-17 11:43 AM
 Received by:
 Relinquished by:

Date/Time 04-11-17 12:40 PM
 Received by:
 Relinquished by: