



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB86815

Ambient Barometric Pressure (mmHg): 760

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | Dilution | Celsius °C | Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----------|------------|-----------|------------|-----------|
| 1   | ICV         | 81            | 1        | 27.0       | 80.6      | 04/11/2017 | 10:00     |
| 2   | I2595-08    |               | 1        | >100.0     | >212.0    | 04/11/2017 | 10:20     |
| 3   | I2595-08DUP |               | 1        | >100.0     | >212.0    | 04/11/2017 | 10:45     |
| 4   | I2619-02    |               | 1        | >100.0     | >212.0    | 04/11/2017 | 11:15     |
| 5   | I2621-02    |               | 1        | >100.0     | >212.0    | 04/11/2017 | 11:45     |

Result = (Celsius \* 1.8) + 32

# WORKLIST(Hardcopy Internal Chain)

LB 86815

WorkList Name : FP41117

WorkList ID : 97332

Date : 4/11/2017 9:33:56 AM

| Due Date   | Matrix | Sample     | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|------------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 04/14/2017 | Water  | I2595-08 C | Flash Point |              | PSEG01   |                  | A2394           | 04/07/2017   | 1010A  |
| 04/12/2017 | Water  | I2619-02 C | Flash Point | Cool 4 deg C | PSEG01   | O41              | SUMP            | 04/10/2017   | 1010A  |
|            | Water  | I2621-02 ✓ | Flash Point | Cool 4 deg C | SEAC01   | O23              | GROUND-WATER    | 04/10/2017   | 1010A  |

Date/Time 4/11/17 9:40 AM  
 Received by: KS  
 Relinquished by: CP

Date/Time 4/11/17 11:50 AM  
 Received by: CP  
 Relinquished by: KS