


7471A

LB86837
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41643



A= 0.0000e+000

B= 2.4152e-004

C= 2.7594e-002

Rho= 0.9998832

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	30	0.000	30				
0.2	0.200	0.234	0.034	853	0.0 %	853				
2.5	2.500	2.460	-0.040	10072	0.0 %	10072				
5.0	5.000	4.970	-0.030	20463	0.0 %	20463				
7.5	7.500	7.416	-0.084	30593	0.0 %	30593				
10.0	10.000	10.085	0.085	41643	0.0 %	41643				

LB86837 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 12 Apr 2017 09:20:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	30	-	0.0000	7471A	PPB	12 Apr 2017 09:25:43	S	Std
0.2 - 1	SO-2	853	-	0.2000	7471A	PPB	12 Apr 2017 09:28:00	S	Std
2.5 - 1	SO-5	10072	-	2.5000	7471A	PPB	12 Apr 2017 09:30:18	S	Std
5.0 - 1	SO	20463	-	5.0000	7471A	PPB	12 Apr 2017 09:32:36	S	Std
7.5 - 1	SO-5	30593	-	7.5000	7471A	PPB	12 Apr 2017 09:34:53	S	Std
10.0 - 1	SO	41643	-	10.0000	7471A	PPB	12 Apr 2017 09:40:19	S	Std
ICV15 - 1	ICV15	16931	4.1168	-	7471A	PPB	12 Apr 2017 09:49:17	U	SMPL
ICB15 - 1	ICB15	-20	0.0228	-	7471A	PPB	12 Apr 2017 09:51:33	U	SMPL
CCV96 - 1	CCV96	20940	5.0850	-	7471A	PPB	12 Apr 2017 09:53:48	U	SMPL
CCB96 - 1	CCB96	-22	0.0223	-	7471A	PPB	12 Apr 2017 09:56:04	U	SMPL
CRA - 1	CRA	772	0.2140	-	7471A	PPB	12 Apr 2017 10:03:08	U	SMPL
HighStd - 1	HighStd	42151	10.2079	-	7471A	PPB	12 Apr 2017 10:05:23	U	SMPL
ChkStd - 1	ChkStd	27754	6.7307	-	7471A	PPB	12 Apr 2017 10:07:40	U	SMPL
PB98208BL - 1	PBS	-31	0.0201	-	7471A	PPB	12 Apr 2017 10:09:57	U	SMPL
PB98208BS - 1	LCSS	14552	3.5422	-	7471A	PPB	12 Apr 2017 10:12:13	U	SMPL
I2611-01 - 1	VIRGINROCK-1	588	0.1696	-	7471A	PPB	12 Apr 2017 10:14:30	U	SMPL
I2620-01 - 1	D001-D002-D003	841	0.2307	-	7471A	PPB	12 Apr 2017 10:16:48	U	SMPL
I2620-06 - 1	D004-D005-D006	1134	0.3015	-	7471A	PPB	12 Apr 2017 10:19:05	U	SMPL
I2620-11 - 1	SOUTH-MAHWAH	1018	0.2735	-	7471A	PPB	12 Apr 2017 10:21:22	U	SMPL
I2622-01 - 1	E2-CONCRETE-BATCH-4	68860	16.6586	-	7471A	PPB	12 Apr 2017 10:23:38	U	SMPL
CCV97 - 1	CCV97	21355	5.1852	-	7471A	PPB	12 Apr 2017 10:25:54	U	SMPL
CCB97 - 1	CCB97	-6	0.0261	-	7471A	PPB	12 Apr 2017 10:28:11	U	SMPL
I2622-03 - 1	E2-CONCRETE-BATCH-5	1463	0.3809	-	7471A	PPB	12 Apr 2017 10:30:27	U	SMPL
I2622-05 - 1	E2-AA8-BASE	47820	11.5770	-	7471A	PPB	12 Apr 2017 10:52:11	U	SMPL
I2622-05DUP - 1	E2-AA8-BASEDUP	47465	11.4913	-	7471A	PPB	12 Apr 2017 10:54:26	U	SMPL
I2622-06 - 1	I2622-05MS	65129	15.7575	-	7471A	PPB	12 Apr 2017 10:59:30	U	SMPL
I2622-07 - 1	I2622-05MSD	65982	15.9635	-	7471A	PPB	12 Apr 2017 11:01:46	U	SMPL
I2622-08 - 1	E2-AA8-BASE-D	217838	52.6397	-	7471A	PPB	12 Apr 2017 11:04:04	U	SMPL
I2622-09 - 1	E2-AA10-WSW	46824	11.3365	-	7471A	PPB	12 Apr 2017 11:06:22	U	SMPL
I2624-01 - 1	IM-TS-WSG-04	3241	0.8104	-	7471A	PPB	12 Apr 2017 11:08:42	U	SMPL
I2625-01 - 1	IM-TS-EME-07	1120	0.2981	-	7471A	PPB	12 Apr 2017 11:11:00	U	SMPL
I2622-01DLX5 - 1		14307	3.4830	-	7471A	PPB	12 Apr 2017 11:13:16	U	SMPL
CCV98 - 1	CCV98	21049	5.1113	-	7471A	PPB	12 Apr 2017 11:15:32	U	SMPL
CCB98 - 1	CCB98	-93	0.0051	-	7471A	PPB	12 Apr 2017 11:17:49	U	SMPL
I2622-05LX5 - 1		9327	2.2802	-	7471A	PPB	12 Apr 2017 11:20:05	U	SMPL
I2622-05A - 1		64681	15.6493	-	7471A	PPB	12 Apr 2017 11:25:07	U	SMPL
I2622-05DLX5 - 1		10111	2.4696	-	7471A	PPB	12 Apr 2017 11:27:23	U	SMPL
I2622-05DUPDLX5 - 1		10150	2.4790	-	7471A	PPB	12 Apr 2017 11:29:40	U	SMPL
I2622-06DLX5 - 1		13096	3.1905	-	7471A	PPB	12 Apr 2017 11:31:57	U	SMPL
I2622-07DLX5 - 1		13438	3.2731	-	7471A	PPB	12 Apr 2017 11:34:14	U	SMPL
I2622-05LDLX25 - 1		1867	0.4785	-	7471A	PPB	12 Apr 2017 11:36:31	U	SMPL
I2622-05ADLX5 - 1		13839	3.3700	-	7471A	PPB	12 Apr 2017 11:38:49	U	SMPL
I2622-08DLX10 - 1		24156	5.8617	-	7471A	PPB	12 Apr 2017 11:41:07	U	SMPL
I2622-09DLX5 - 1		9128	2.2322	-	7471A	PPB	12 Apr 2017 11:43:23	U	SMPL
CCV99 - 1	CCV99	21010	5.1019	-	7471A	PPB	12 Apr 2017 11:45:39	U	SMPL
CCB99 - 1	CCB99	19	0.0322	-	7471A	PPB	12 Apr 2017 11:47:56	U	SMPL