


7470A

LB86839
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37638



A= 0.0000e+000

B= 2.6391e-004

C= -4.6933e-002

Rho= 0.9997815

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	99	0.000	99				
0.2	0.200	0.176	-0.024	844	0.0 %	844				
2.5	2.500	2.520	0.020	9727	0.0 %	9727				
5.0	5.000	4.993	-0.007	19098	0.0 %	19098				
7.5	7.500	7.646	0.146	29149	0.0 %	29149				
10.0	10.000	9.886	-0.114	37638	0.0 %	37638				

LB86839 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Apr 2017 12:06:01

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	99	-	0.0000	7470A	PPB	12 Apr 2017 12:08:07	S	Std
0.2 - 1	50.2	844	-	0.2000	7470A	PPB	12 Apr 2017 12:10:15	S	Std
2.5 - 1	52.5	9727	-	2.5000	7470A	PPB	12 Apr 2017 12:12:23	S	Std
5.0 - 1	55	19098	-	5.0000	7470A	PPB	12 Apr 2017 12:14:31	S	Std
7.5 - 1	57.5	29149	-	7.5000	7470A	PPB	12 Apr 2017 12:16:39	S	Std
10.0 - 1	510	37638	-	10.0000	7470A	PPB	12 Apr 2017 12:18:46	S	Std
ICV16 - 1	ICV16	14338	3.7370	-	7470A	PPB	12 Apr 2017 12:21:31	U	SMPL
ICB16 - 1	ICB16	60	-0.0311	-	7470A	PPB	12 Apr 2017 12:23:37	U	SMPL
CCV01 - 1	CCV01	17604	4.5989	-	7470A	PPB	12 Apr 2017 12:25:43	U	SMPL
CCB01 - 1	CCB01	54	-0.0327	-	7470A	PPB	12 Apr 2017 12:27:50	U	SMPL
CRA - 1	CRA	815	0.1682	-	7470A	PPB	12 Apr 2017 12:29:56	U	SMPL
HighStd - 1	HighStd	37425	9.8298	-	7470A	PPB	12 Apr 2017 12:32:03	U	SMPL
ChkStd - 1	ChkStd	26716	7.0036	-	7470A	PPB	12 Apr 2017 12:34:10	U	SMPL
PB98199BL - 1	PBW	55	-0.0324	-	7470A	PPB	12 Apr 2017 12:36:18	U	SMPL
PB98199BS - 1	LCSW	14506	3.7813	-	7470A	PPB	12 Apr 2017 12:38:25	U	SMPL
I2595-07 - 1	A2394	245	0.0177	-	7470A	PPB	12 Apr 2017 12:40:33	U	SMPL
I2602-01 - 1	JAN-GW-SP4-30-34	1064	0.2339	-	7470A	PPB	12 Apr 2017 12:47:14	U	SMPL
I2602-01DUP - 1	JAN-GW-SP4-30-34DUP	1032	0.2254	-	7470A	PPB	12 Apr 2017 12:49:21	U	SMPL
I2602-01MS - 1	JAN-GW-SP4-30-34MS	16069	4.1938	-	7470A	PPB	12 Apr 2017 12:51:30	U	SMPL
I2602-01MSD - 1	JAN-GW-SP4-30-34MSD	16214	4.2321	-	7470A	PPB	12 Apr 2017 12:53:37	U	SMPL
CCV02 - 1	CCV02	18037	4.7132	-	7470A	PPB	12 Apr 2017 12:55:43	U	SMPL
CCB02 - 1	CCB02	65	-0.0298	-	7470A	PPB	12 Apr 2017 12:57:50	U	SMPL
I2602-02 - 1	JAN-GW-SP9-30-34	537	0.0948	-	7470A	PPB	12 Apr 2017 12:59:57	U	SMPL
I2602-03 - 1	JAN-GW-SP9-40-44	431	0.0668	-	7470A	PPB	12 Apr 2017 13:02:03	U	SMPL
I2602-04 - 1	FB-2017-04-06	141	-0.0097	-	7470A	PPB	12 Apr 2017 13:04:09	U	SMPL
I2602-05 - 1	FB-2017-04-07	132	-0.0121	-	7470A	PPB	12 Apr 2017 13:06:16	U	SMPL
I2602-06 - 1	JAN-GW-SP9-80-84	2788	0.6888	-	7470A	PPB	12 Apr 2017 13:08:24	U	SMPL
I2602-07 - 1	JAN-GW-SP9-80-84-DUP	2061	0.4970	-	7470A	PPB	12 Apr 2017 13:10:31	U	SMPL
I2602-11 - 1	JAN-GW-SP4-20-24	1582	0.3706	-	7470A	PPB	12 Apr 2017 13:12:39	U	SMPL
I2619-01 - 1	SUMP	127	-0.0134	-	7470A	PPB	12 Apr 2017 13:14:48	U	SMPL
I2621-02 - 1	GROUND-WATER	354	0.0465	-	7470A	PPB	12 Apr 2017 13:16:57	U	SMPL
I2622-10 - 1	E2-FB-4-10	130	-0.0126	-	7470A	PPB	12 Apr 2017 13:19:04	U	SMPL
CCV03 - 1	CCV03	18089	4.7269	-	7470A	PPB	12 Apr 2017 13:21:10	U	SMPL
CCB03 - 1	CCB03	87	-0.0240	-	7470A	PPB	12 Apr 2017 13:23:17	U	SMPL
PB98200BL - 1	PBW	187	0.0024	-	7470A	PPB	12 Apr 2017 13:25:24	U	SMPL
PB98200BS - 1	LCSW	14303	3.7277	-	7470A	PPB	12 Apr 2017 13:27:31	U	SMPL
I2603-01 - 1	JAN-GW-SP7-20-24	127618	33.6324	-	7470A	PPB	12 Apr 2017 13:29:38	U	SMPL
I2603-01DUP - 1	JAN-GW-SP7-20-24DUP	123330	32.5008	-	7470A	PPB	12 Apr 2017 13:31:45	U	SMPL
I2603-01MS - 1	JAN-GW-SP7-20-24MS	131281	34.5991	-	7470A	PPB	12 Apr 2017 13:33:54	U	SMPL
I2603-01MSD - 1	JAN-GW-SP7-20-24MSD	126396	33.3099	-	7470A	PPB	12 Apr 2017 13:36:04	U	SMPL
I2603-02 - 1	JAN-GW-SP7-30-34	3061	0.7609	-	7470A	PPB	12 Apr 2017 13:38:13	U	SMPL
I2603-03 - 1	JAN-GW-SP7-40-44	14787	3.8555	-	7470A	PPB	12 Apr 2017 13:40:23	U	SMPL
I2603-04 - 1	JAN-GW-SP7-60-64	9596	2.4855	-	7470A	PPB	12 Apr 2017 13:42:32	U	SMPL
I2603-05 - 1	JAN-GW-SP8-80-84	292	0.0301	-	7470A	PPB	12 Apr 2017 13:44:39	U	SMPL
CCV04 - 1	CCV04	18088	4.7261	-	7470A	PPB	12 Apr 2017 13:46:46	U	SMPL
CCB04 - 1	CCB04	-69	-0.0651	-	7470A	PPB	12 Apr 2017 13:48:53	U	SMPL
I2603-06 - 1	JAN-GW-SP8-70-74	609	0.1138	-	7470A	PPB	12 Apr 2017 13:51:00	U	SMPL
I2603-07 - 1	JAN-GW-SP8-60-64	383	0.0541	-	7470A	PPB	12 Apr 2017 13:53:08	U	SMPL
I2603-08 - 1	JAN-GW-SP8-50-54	425	0.0652	-	7470A	PPB	12 Apr 2017 13:55:15	U	SMPL
I2603-09 - 1	JAN-GW-SP8-40-44	664	0.1283	-	7470A	PPB	12 Apr 2017 13:57:22	U	SMPL
I2603-10 - 1	JAN-GW-SP8-30-34	407	0.0605	-	7470A	PPB	12 Apr 2017 13:59:29	U	SMPL
I2603-11 - 1	JAN-GW-SP6-20-24	1336	0.3056	-	7470A	PPB	12 Apr 2017 14:01:37	U	SMPL
I2603-12 - 1	JAN-GW-SP6-30-34	2660	0.6551	-	7470A	PPB	12 Apr 2017 14:03:45	U	SMPL
I2603-13 - 1	JAN-GW-SP6-50-54	575	0.1048	-	7470A	PPB	12 Apr 2017 14:05:53	U	SMPL
I2603-14 - 1	JAN-GW-SP6-50-54-DUP	507	0.0869	-	7470A	PPB	12 Apr 2017 14:08:02	U	SMPL
I2603-15 - 1	FB-2017-04-04	107	-0.0187	-	7470A	PPB	12 Apr 2017 14:10:09	U	SMPL
CCV05 - 1	CCV05	17759	4.6398	-	7470A	PPB	12 Apr 2017 14:12:16	U	SMPL
CCB05 - 1	CCB05	-42	-0.0580	-	7470A	PPB	12 Apr 2017 14:14:23	U	SMPL
I2603-16 - 1	JAN-GW-SP6-60-64	1543	0.3603	-	7470A	PPB	12 Apr 2017 14:16:30	U	SMPL
I2603-17 - 1	JAN-GW-SP6-70-74	453	0.0726	-	7470A	PPB	12 Apr 2017 14:18:37	U	SMPL
I2603-18 - 1	JAN-GW-SP6-80-84	10111	2.6214	-	7470A	PPB	12 Apr 2017 14:20:44	U	SMPL
I2603-20 - 1	FB-2017-04-05	78	-0.0263	-	7470A	PPB	12 Apr 2017 14:22:52	U	SMPL
PB98207BL - 1	PBW	53	-0.0329	-	7470A	PPB	12 Apr 2017 14:24:59	U	SMPL
PB98207BS - 1	LCSW	14089	3.6713	-	7470A	PPB	12 Apr 2017 14:29:51	U	SMPL
PB98207TB - 1	PB98174TB	0	-0.0469	-	7470A	PPB	12 Apr 2017 14:31:59	U	SMPL
I2612-03 - 1	HOIL-3	-7	-0.0488	-	7470A	PPB	12 Apr 2017 14:34:07	U	SMPL
I2619-02 - 1	SUMP	33	-0.0382	-	7470A	PPB	12 Apr 2017 14:36:16	U	SMPL
I2620-05 - 1	D001-D002-D003	89	-0.0234	-	7470A	PPB	12 Apr 2017 14:38:23	U	SMPL
CCV06 - 1	CCV06	18002	4.7039	-	7470A	PPB	12 Apr 2017 14:40:30	U	SMPL
CCB06 - 1	CCB06	-26	-0.0538	-	7470A	PPB	12 Apr 2017 14:42:38	U	SMPL
I2620-10 - 1	D004-D005-D006	60	-0.0311	-	7470A	PPB	12 Apr 2017 14:44:45	U	SMPL
I2620-15 - 1	SOUTH-MAHWAH	118	-0.0158	-	7470A	PPB	12 Apr 2017 14:46:52	U	SMPL
I2620-15DUP - 1	SOUTH-MAHWAHDUP	212	0.0090	-	7470A	PPB	12 Apr 2017 14:49:00	U	SMPL
I2620-15MS - 1	SOUTH-MAHWAHMS	14519	3.7847	-	7470A	PPB	12 Apr 2017 14:51:07	U	SMPL
I2620-15MSD - 1	SOUTH-MAHWAHMSD	13539	3.5261	-	7470A	PPB	12 Apr 2017 14:53:14	U	SMPL
I2621-01 - 1	SOIL-CUTTINGS	43	-0.0356	-	7470A	PPB	12 Apr 2017 14:55:22	U	SMPL
I2602-01LX5 - 1		525	0.0916	-	7470A	PPB	12 Apr 2017 14:57:30	U	SMPL
I2602-01A - 1		17229	4.4999	-	7470A	PPB	12 Apr 2017 14:59:39	U	SMPL
I2603-01LX5 - 1		23726	6.2145	-	7470A	PPB	12 Apr 2017 15:01:47	U	SMPL
I2603-01A - 1		145221	38.2780	-	7470A	PPB	12 Apr 2017 15:07:32	U	SMPL
CCV07 - 1	CCV07	18402	4.8095	-	7470A	PPB	12 Apr 2017 15:09:39	U	SMPL
CCB07 - 1	CCB07	-44	-0.0585	-	7470A	PPB	12 Apr 2017 15:11:48	U	SMPL
I2620-15LX5 - 1		69	-0.0287	-	7470A	PPB	12 Apr 2017 15:13:56	U	SMPL
I2620-15A - 1		14906	3.8869	-	7470A	PPB	12 Apr 2017 15:16:03	U	SMPL
I2603-01DLX5 - 1		24058	6.3022	-	7470A	PPB	12 Apr 2017 15:18:11	U	SMPL
I2603-01DUPDLX5 - 1		22981	6.0179	-	7470A	PPB	12 Apr 2017 15:20:19	U	SMPL
I2603-01MSDLX5 - 1		24120	6.3185	-	7470A	PPB	12 Apr 2017 15:22:27	U	SMPL
I2603-01MSDDLX5 - 1		23263	6.0923	-	7470A	PPB	12 Apr 2017 15:24:34	U	SMPL
I2603-01LDLX25 - 1		4621	1.1726	-	7470A	PPB	12 Apr 2017 15:26:43	U	SMPL
I2603-01ADLX5 - 1		26531	6.9548	-	7470A	PPB	12 Apr 2017 15:28:51	U	SMPL
CCV08 - 1	CCV08	18325	4.7892	-	7470A	PPB	12 Apr 2017 15:31:00	U	SMPL
CCB08 - 1	CCB08	7	-0.0451	-	7470A	PPB	12 Apr 2017 15:33:08	U	SMPL