

245.1


LB86842

INSTRUMENT ID: CY1

Linear

 μ Abs.:

34453



A= 0.0000e+000

B= 2.9233e-004

C= -1.7109e-002

Rho= 0.9999274

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	68	0.000	68				
0.2	0.200	0.197	-0.003	731	0.0 %	731				
2.5	2.500	2.525	0.025	8695	0.0 %	8695				
5.0	5.000	5.012	0.012	17202	0.0 %	17202				
7.5	7.500	7.410	-0.090	25406	0.0 %	25406				
10.0	10.000	10.055	0.055	34453	0.0 %	34453				

LB86842 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Apr 2017 16:03:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	68	-	0.0000	245.1	PPB	12 Apr 2017 16:05:55	S	Std
0.2 - 1	50.2	731	-	0.2000	245.1	PPB	12 Apr 2017 16:08:03	S	Std
2.5 - 1	52.5	8695	-	2.5000	245.1	PPB	12 Apr 2017 16:10:12	S	Std
5.0 - 1	55	17202	-	5.0000	245.1	PPB	12 Apr 2017 16:12:20	S	Std
7.5 - 1	57.5	25406	-	7.5000	245.1	PPB	12 Apr 2017 16:14:28	S	Std
10.0 - 1	510	34453	-	10.0000	245.1	PPB	12 Apr 2017 16:16:35	S	Std
ICV17 - 1	ICV17	13638	3.9697	-245.1		PPB	12 Apr 2017 16:19:05	U	SMPL
ICB17 - 1	ICB17	0	-0.0171	-245.1		PPB	12 Apr 2017 16:21:11	U	SMPL
CCV09 - 1	CCV09	17058	4.9695	-245.1		PPB	12 Apr 2017 16:23:17	U	SMPL
CCB09 - 1	CCB09	-9	-0.0197	-245.1		PPB	12 Apr 2017 16:25:24	U	SMPL
CRA - 1	CRA	709	0.1902	-245.1		PPB	12 Apr 2017 16:27:31	U	SMPL
HighStd - 1	HighStd	34134	9.9613	-245.1		PPB	12 Apr 2017 16:29:37	U	SMPL
ChkStd - 1	ChkStd	24071	7.0196	-245.1		PPB	12 Apr 2017 16:31:44	U	SMPL
PB98198BL - 1	PBW	41	-0.0051	-245.1		PPB	12 Apr 2017 16:33:52	U	SMPL
PB98198BS - 1	LCSW	14572	4.2427	-245.1		PPB	12 Apr 2017 16:36:00	U	SMPL
I2602-10 - 1	256-TUCKAHOE	59	0.0001	-245.1		PPB	12 Apr 2017 16:38:08	U	SMPL
I2602-10DUP - 1	256-TUCKAHOEDUP	123	0.0188	-245.1		PPB	12 Apr 2017 16:40:16	U	SMPL
I2602-10MS - 1	256-TUCKAHOEMS	14105	4.1062	-245.1		PPB	12 Apr 2017 16:42:24	U	SMPL
I2602-10MSD - 1	256-TUCKAHOEMSD	14213	4.1378	-245.1		PPB	12 Apr 2017 16:44:33	U	SMPL
CCV10 - 1	CCV10	17475	5.0914	-245.1		PPB	12 Apr 2017 16:46:39	U	SMPL
CCB10 - 1	CCB10	-15	-0.0215	-245.1		PPB	12 Apr 2017 16:48:46	U	SMPL
I2602-10LX5 - 1		65	0.0019	-245.1		PPB	12 Apr 2017 16:50:53	U	SMPL
I2602-10A - 1		14147	4.1185	-245.1		PPB	12 Apr 2017 16:52:59	U	SMPL
CCV11 - 1	CCV11	17361	5.0580	-245.1		PPB	12 Apr 2017 16:55:06	U	SMPL
CCB11 - 1	CCB11	-23	-0.0238	-245.1		PPB	12 Apr 2017 16:57:13	U	SMPL