



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86846

Temp.: 23.2 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/12/2017	17:20
2	CAL2	1	Water			7.01	04/12/2017	17:21
3	CAL3	1	Water			10.03	04/12/2017	17:22
4	ICV	1	Water			7.01	04/12/2017	17:23
5	CCV1	1	Water			2.02	04/12/2017	17:25
6	I2662-09	1	Solid	20.03	20	9.10	04/12/2017	17:26
7	I2662-09DUP	1	Solid	20.04	20	9.11	04/12/2017	17:27
8	I2662-10	1	Solid	20.05	20	9.25	04/12/2017	17:28
9	I2662-11	1	Solid	20.02	20	9.14	04/12/2017	17:30
10	I2662-12	1	Solid	20.01	20	10.51	04/12/2017	17:31
11	I2670-03	1	Solid	20.02	20	7.35	04/12/2017	17:32
12	CCV2	1	Water			12.02	04/12/2017	17:33

2886846

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2670 WorkList ID : 97363 Date : 4/12/2017 4:57:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2662-09	Corrosivity	Cool 4 deg C	TULL01	O63	AREA-1-COMP	04/12/2017	9045C
	Solid	I2662-10	Corrosivity	Cool 4 deg C	TULL01	O63	AREA-2-COMP	04/12/2017	9045C
	Solid	I2662-11	Corrosivity	Cool 4 deg C	TULL01	O63	AREA-3-COMP	04/12/2017	9045C
	Solid	I2662-12	Corrosivity	Cool 4 deg C	TULL01	O63	AREA-4-COMP	04/12/2017	9045C
04/17/2017	Solid	I2670-03	Corrosivity	Cool 4 deg C	PSEG01	O51	ROLL-OFF-R4	04/12/2017	9045C

Date/Time 04-12-17 4:57 PM
 Received by: SP
 Relinquished by: SP

Date/Time 04-12-17 6:50 PM
 Received by: CR
 Relinquished by: SP