



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86847

Temp.: 23.2 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
ZOBELLS SOLUTION ORP STD 500ML	W2118
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/12/2017	17:20
2	CAL2	1	Water			7.01	04/12/2017	17:21
3	CAL3	1	Water			10.03	04/12/2017	17:22
4	ICV	1	Water			7.01	04/12/2017	17:23
5	CCV1	1	Water			2.01	04/12/2017	17:40
6	I2671-03	1	Water			6.56	04/12/2017	17:41
7	I2671-03DUP	1	Water			6.57	04/12/2017	17:42
8	I2671-06	1	Water			5.01	04/12/2017	17:43
9	CCV2	1	Water			12.02	04/12/2017	17:44

WORKLIST(Hardcopy Internal Chain)

1386847

WorkList Name : ph i2671

WorkList ID : 97364

Date : 4/12/2017 4:57:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/12/2017	Water	I2671-03	pH	Cool 4 deg C	PSEG01	O52	DRUMS-(2-7)	04/12/2017	9045C
04/12/2017	Water	I2671-06	pH	Cool 4 deg C	PSEG01	O52	M-15-M-18	04/12/2017	9045C

Date/Time 04-12-17 11:37 PM
 Received by: JP
 Relinquished by: JP

Date/Time 04-12-17 7:00 AM
 Received by: CP
 Relinquished by: JP