



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86848

Temp.: 23.2 °C

BalanceID: WC SC-3

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/12/2017	17:20
2	CAL2	1	Water			7.01	04/12/2017	17:21
3	CAL3	1	Water			10.03	04/12/2017	17:22
4	ICV	1	Water			7.01	04/12/2017	17:23
5	CCV1	1	Water			2.02	04/12/2017	17:50
6	I2663-01	1	Solid	20.03	20	7.22	04/12/2017	17:51
7	I2663-01DUP	1	Solid	20.04	20	7.23	04/12/2017	17:52
8	I2666-01	1	Solid	20.05	20	8.41	04/12/2017	17:53
9	CCV2	1	Water			12.02	04/12/2017	17:54

1086848

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 12663

WorkList ID : 97365

Date : 4/12/2017 4:58:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2663-01	pH	Cool 4 deg C	TULL02	O51	EME-BW-PLANTINGSOIL	04/11/2017	9045C
	Solid	I2666-01	pH	Cool 4 deg C	TULL02	O43	EME-PRODUCED-4-10	04/11/2017	9045C

Date/Time 04-12-17 4:58 PM
Received by: J
Relinquished by: J

Date/Time 04-12-17 6:50 PM
Received by: CP
Relinquished by: J