



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB86875

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	04/13/2017	15:30
2	I2671-04		1	>100.0	>212.0	04/13/2017	16:00
3	I2671-04DUP		1	>100.0	>212.0	04/14/2017	16:30
4	I2671-07		1	>100.0	>212.0	04/14/2017	16:50

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 86875

Worklist Name : FP41317

Worklist ID : 97415

Date : 4/13/2017 2:51:52 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/17/2017	Water	12671-04 B	Flash Point	Cool 4 deg C	PSEG01	052	DRUMS-(2-7)	04/12/2017	1010A
04/17/2017	Water	12671-07 B	Flash Point	Cool 4 deg C	PSEG01	052	M-15-M-18	04/12/2017	1010A

Date/Time 4/13/17 3:00 PM
 Received by: KS
 Relinquished by: [Signature]

Date/Time 4/13/17 17:00
 Received by: [Signature]
 Relinquished by: KS