


245.1

LB86876
INSTRUMENT ID: CV2

Linear

 μ Abs.:

33364



A= 0.0000e+000

B= 2.9881e-004

C= -6.0195e-003

Rho= 0.9999615

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	42	0.000	42				
0.2	0.200	0.206	0.006	708	0.0 %	708				
2.5	2.500	2.468	-0.032	8280	0.0 %	8280				
5.0	5.000	4.994	-0.006	16733	0.0 %	16733				
7.5	7.500	7.562	0.062	25327	0.0 %	25327				
10.0	10.000	9.964	-0.036	33364	0.0 %	33364				

LB86876 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 14 Apr 2017 08:24:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	42	-	0.0000	245.1	PPB	14 Apr 2017 08:36:53	S	Std
0.2 - 1	50.2	708	-	0.2000	245.1	PPB	14 Apr 2017 08:39:01	S	Std
2.5 - 1	52.5	8280	-	2.5000	245.1	PPB	14 Apr 2017 08:41:10	S	Std
5.0 - 1	55	16733	-	5.0000	245.1	PPB	14 Apr 2017 08:43:18	S	Std
7.5 - 1	57.5	25327	-	7.5000	245.1	PPB	14 Apr 2017 08:45:26	S	Std
10.0 - 1	50	33364	-	10.0000	245.1	PPB	14 Apr 2017 08:50:10	S	Std
ICV18 - 1	ICV18	13799	4.1173	-245.1		PPB	14 Apr 2017 08:52:43	U	SMPL
ICB18 - 1	ICB18	-45	-0.0195	-245.1		PPB	14 Apr 2017 08:54:49	U	SMPL
CCV12 - 1	CCV12	16804	5.0153	-245.1		PPB	14 Apr 2017 08:56:55	U	SMPL
CCB12 - 1	CCB12	-49	-0.0207	-245.1		PPB	14 Apr 2017 08:59:02	U	SMPL
CRA - 1	CRA	701	0.2034	-245.1		PPB	14 Apr 2017 09:01:09	U	SMPL
HighStd - 1	HighStd	34027	10.1618	-245.1		PPB	14 Apr 2017 09:03:15	U	SMPL
ChkStd - 1	ChkStd	23572	7.0376	-245.1		PPB	14 Apr 2017 09:05:22	U	SMPL
PB98242BL - 1	PBW	-79	-0.0296	-245.1		PPB	14 Apr 2017 09:07:31	U	SMPL
PB98242BS - 1	LCSW	13256	3.9551	-245.1		PPB	14 Apr 2017 09:09:39	U	SMPL
I2649-01 - 1	32220008-3	-75	-0.0284	-245.1		PPB	14 Apr 2017 09:11:46	U	SMPL
I2649-01DUP - 1	32220008-3DUP	-15	-0.0105	-245.1		PPB	14 Apr 2017 09:13:55	U	SMPL
I2649-01MS - 1	32220008-3MS	12379	3.6930	-245.1		PPB	14 Apr 2017 09:16:03	U	SMPL
I2649-01MSD - 1	32220008-3MSD	12149	3.6243	-245.1		PPB	14 Apr 2017 09:18:11	U	SMPL
I2677-01 - 1	041117-DSDC-F-D-M	-81	-0.0302	-245.1		PPB	14 Apr 2017 09:20:18	U	SMPL
CCV13 - 1	CCV13	16865	5.0335	-245.1		PPB	14 Apr 2017 09:22:25	U	SMPL
CCB13 - 1	CCB13	-43	-0.0189	-245.1		PPB	14 Apr 2017 09:24:32	U	SMPL
I2677-02 - 1	041117-DSDC-F-U-M	0	-0.0060	-245.1		PPB	14 Apr 2017 09:26:39	U	SMPL
I2677-03 - 1	041117-PRRT-E-D-M	65	0.0134	-245.1		PPB	14 Apr 2017 09:28:45	U	SMPL
I2678-01 - 1	041117-PRRT-E-U-M	5	-0.0045	-245.1		PPB	14 Apr 2017 09:30:52	U	SMPL
I2649-01LX5 - 1		16	-0.0012	-245.1		PPB	14 Apr 2017 09:32:59	U	SMPL
I2649-01A - 1		13663	4.0767	-245.1		PPB	14 Apr 2017 09:35:06	U	SMPL
CCV14 - 1	CCV14	16678	4.9776	-245.1		PPB	14 Apr 2017 09:37:14	U	SMPL
CCB14 - 1	CCB14	-44	-0.0192	-245.1		PPB	14 Apr 2017 09:39:22	U	SMPL