


7470A

LB 8688 U
INSTRUMENT ID: CV 1

Linear

 μ Abs.:

36273



A= 0.0000e+000

B= 2.7496e-004

C= -1.1690e-002

Rho= 0.9999511

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	7	0.000	7				
0.2	0.200	0.176	-0.024	684	0.0 %	684				
2.5	2.500	2.493	-0.007	9111	0.0 %	9111				
5.0	5.000	5.076	0.076	18505	0.0 %	18505				
7.5	7.500	7.501	0.001	27324	0.0 %	27324				
10.0	10.000	9.962	-0.038	36273	0.0 %	36273				

LB86884 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Apr 2017 09:51:36

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	7	-	0.0000	7470A	PPB	14 Apr 2017 10:09:04	S	Std
0.2 - 1	50.2	684	-	0.2000	7470A	PPB	14 Apr 2017 10:11:12	S	Std
2.5 - 1	50.5	9111	-	2.5000	7470A	PPB	14 Apr 2017 10:13:21	S	Std
5.0 - 1	55	18505	-	5.0000	7470A	PPB	14 Apr 2017 10:15:29	S	Std
7.5 - 1	57.5	27324	-	7.5000	7470A	PPB	14 Apr 2017 10:17:37	S	Std
10.0 - 1	510	36273	-	10.0000	7470A	PPB	14 Apr 2017 10:22:20	S	Std
ICV19 - 1	ICV19	14188	3.8895	-	7470A	PPB	14 Apr 2017 10:24:52	U	SMPL
ICB19 - 1	ICB19	-77	-0.0329	-	7470A	PPB	14 Apr 2017 10:26:58	U	SMPL
CCV15 - 1	CCV15	18241	5.0039	-	7470A	PPB	14 Apr 2017 10:29:05	U	SMPL
CCB15 - 1	CCB15	-94	-0.0375	-	7470A	PPB	14 Apr 2017 10:31:11	U	SMPL
CRA - 1	CRA	671	0.1728	-	7470A	PPB	14 Apr 2017 10:33:18	U	SMPL
HighStd - 1	HighStd	36292	9.9672	-	7470A	PPB	14 Apr 2017 10:35:25	U	SMPL
ChkStd - 1	ChkStd	25122	6.8959	-	7470A	PPB	14 Apr 2017 10:37:32	U	SMPL
PB98240BL - 1	PBW	-103	-0.0400	-	7470A	PPB	14 Apr 2017 10:39:39	U	SMPL
PB98240BS - 1	LCSW	13947	3.8232	-	7470A	PPB	14 Apr 2017 10:41:47	U	SMPL
I2653-01 - 1	JAN-GW-SP4-60-64	4421	1.2039	-	7470A	PPB	14 Apr 2017 10:43:55	U	SMPL
I2653-02 - 1	JAN-GW-SP4-80-84	1572	0.4206	-	7470A	PPB	14 Apr 2017 10:46:03	U	SMPL
I2653-03 - 1	JAN-GW-SP1-20-24	658	0.1692	-	7470A	PPB	14 Apr 2017 10:48:11	U	SMPL
I2653-04 - 1	JAN-GW-SP1-30-34	477	0.1195	-	7470A	PPB	14 Apr 2017 10:50:20	U	SMPL
I2653-05 - 1	JAN-GW-SP1-30-34-DUP	1070	0.2825	-	7470A	PPB	14 Apr 2017 10:52:26	U	SMPL
CCV16 - 1	CCV16	18383	5.0429	-	7470A	PPB	14 Apr 2017 10:54:33	U	SMPL
CCB16 - 1	CCB16	-28	-0.0194	-	7470A	PPB	14 Apr 2017 10:56:40	U	SMPL
I2653-06 - 1	FB-2017-04-10	10	-0.0089	-	7470A	PPB	14 Apr 2017 10:58:47	U	SMPL
I2653-07 - 1	JAN-GW-SP1-40-44	3083	0.8360	-	7470A	PPB	14 Apr 2017 11:00:54	U	SMPL
I2653-08 - 1	JAN-GW-SP1-85-89	19336	5.3050	-	7470A	PPB	14 Apr 2017 11:03:01	U	SMPL
I2653-09 - 1	JAN-GW-SP2-20-24	6089	1.6626	-	7470A	PPB	14 Apr 2017 11:05:08	U	SMPL
I2653-10 - 1	JAN-GW-SP2-30-34	2152	0.5800	-	7470A	PPB	14 Apr 2017 11:07:16	U	SMPL
I2671-03 - 1	DRUMS-(2-7)	58	0.0043	-	7470A	PPB	14 Apr 2017 11:09:23	U	SMPL
I2671-03DUP - 1	DRUMS-(2-7)DUP	46	0.0010	-	7470A	PPB	14 Apr 2017 11:11:31	U	SMPL
I2671-03MS - 1	DRUMS-(2-7)MS	13833	3.7919	-	7470A	PPB	14 Apr 2017 11:13:39	U	SMPL
I2671-03MSD - 1	DRUMS-(2-7)MSD	13926	3.8174	-	7470A	PPB	14 Apr 2017 11:15:48	U	SMPL
I2671-06 - 1	M-15-M-18	-2	-0.0122	-	7470A	PPB	14 Apr 2017 11:17:55	U	SMPL
CCV17 - 1	CCV17	18512	5.0784	-	7470A	PPB	14 Apr 2017 11:20:02	U	SMPL
CCB17 - 1	CCB17	-34	-0.0210	-	7470A	PPB	14 Apr 2017 11:22:09	U	SMPL
PB98241BL - 1	PBW	15	-0.0076	-	7470A	PPB	14 Apr 2017 11:24:16	U	SMPL
PB98241BS - 1	LCSW	13862	3.7998	-	7470A	PPB	14 Apr 2017 11:26:23	U	SMPL
I2655-01 - 1	JAN-GW-SP2-40-44	16713	4.5838	-	7470A	PPB	14 Apr 2017 11:28:30	U	SMPL
I2655-02 - 1	JAN-GW-SP2-55-59	6320	1.7261	-	7470A	PPB	14 Apr 2017 11:30:37	U	SMPL
I2655-03 - 1	JAN-GW-SP3-20-24	2587	0.6996	-	7470A	PPB	14 Apr 2017 11:32:44	U	SMPL
I2655-04 - 1	JAN-GW-SP3-40-44	4420	1.2036	-	7470A	PPB	14 Apr 2017 11:34:51	U	SMPL
I2655-05 - 1	FB-2017-04-11	9	-0.0092	-	7470A	PPB	14 Apr 2017 11:37:00	U	SMPL
I2655-06 - 1	JAN-GW-SP3-60-64	2664	0.7208	-	7470A	PPB	14 Apr 2017 11:39:08	U	SMPL
I2655-07 - 1	JAN-GW-SP5-20-24	2776	0.7516	-	7470A	PPB	14 Apr 2017 11:41:16	U	SMPL
I2655-08 - 1	JAN-GW-SP5-30-34	364	0.0884	-	7470A	PPB	14 Apr 2017 11:43:23	U	SMPL
CCV18 - 1	CCV18	18219	4.9978	-	7470A	PPB	14 Apr 2017 11:45:31	U	SMPL
CCB18 - 1	CCB18	-22	-0.0177	-	7470A	PPB	14 Apr 2017 11:47:38	U	SMPL
I2655-09 - 1	JAN-GW-SP5-40-44	2340	0.6317	-	7470A	PPB	14 Apr 2017 11:49:45	U	SMPL
I2655-10 - 1	JAN-GW-SP5-60-64	1039	0.2740	-	7470A	PPB	14 Apr 2017 11:51:52	U	SMPL
I2655-13 - 1	JAN-GW-SP5-80-84	1214	0.3221	-	7470A	PPB	14 Apr 2017 11:53:59	U	SMPL
I2655-13DUP - 1	JAN-GW-SP5-80-84DUP	1147	0.3037	-	7470A	PPB	14 Apr 2017 11:56:07	U	SMPL
I2655-13MS - 1	JAN-GW-SP5-80-84MS	17106	4.6918	-	7470A	PPB	14 Apr 2017 11:58:14	U	SMPL
I2655-13MSD - 1	JAN-GW-SP5-80-84MSD	16462	4.5147	-	7470A	PPB	14 Apr 2017 12:00:22	U	SMPL
PB98243BL - 1	PBW	0	-0.0117	-	7470A	PPB	14 Apr 2017 12:02:30	U	SMPL
PB98243BS - 1	LCSW	13987	3.8342	-	7470A	PPB	14 Apr 2017 12:04:38	U	SMPL
PB98243TB - 1	PB98213TB	-13	-0.0153	-	7470A	PPB	14 Apr 2017 12:06:47	U	SMPL
I2662-09 - 1	AREA-1-COMP	14	-0.0078	-	7470A	PPB	14 Apr 2017 12:08:54	U	SMPL
CCV19 - 1	CCV19	18204	4.9937	-	7470A	PPB	14 Apr 2017 12:11:02	U	SMPL
CCB19 - 1	CCB19	-36	-0.0216	-	7470A	PPB	14 Apr 2017 12:13:09	U	SMPL
I2662-10 - 1	AREA-2-COMP	72	0.0081	-	7470A	PPB	14 Apr 2017 12:15:16	U	SMPL
I2662-11 - 1	AREA-3-COMP	15	-0.0076	-	7470A	PPB	14 Apr 2017 12:17:23	U	SMPL
I2662-12 - 1	AREA-4-COMP	13	-0.0081	-	7470A	PPB	14 Apr 2017 12:19:30	U	SMPL
I2670-03 - 1	ROLL-OFF-R4	22	-0.0056	-	7470A	PPB	14 Apr 2017 12:21:37	U	SMPL
I2670-03DUP - 1	ROLL-OFF-R4DUP	26	-0.0045	-	7470A	PPB	14 Apr 2017 12:23:45	U	SMPL
I2670-03MS - 1	ROLL-OFF-R4MS	13487	3.6967	-	7470A	PPB	14 Apr 2017 12:25:52	U	SMPL
I2670-03MSD - 1	ROLL-OFF-R4MSD	13005	3.5642	-	7470A	PPB	14 Apr 2017 12:28:00	U	SMPL
I2671-02 - 1	DRUM-(1)	-38	-0.0221	-	7470A	PPB	14 Apr 2017 12:30:09	U	SMPL
I2671-04 - 1	DRUMS-(2-7)	0	-0.0117	-	7470A	PPB	14 Apr 2017 12:32:17	U	SMPL
I2671-07 - 1	M-15-M-18	5	-0.0103	-	7470A	PPB	14 Apr 2017 12:34:24	U	SMPL
CCV20 - 1	CCV20	18654	5.1175	-	7470A	PPB	14 Apr 2017 12:36:31	U	SMPL
CCB20 - 1	CCB20	-58	-0.0276	-	7470A	PPB	14 Apr 2017 12:38:39	U	SMPL
2671-03LX5 - 1		98	0.0153	-	7470A	PPB	14 Apr 2017 12:40:46	U	SMPL
2671-03A - 1		14164	3.8829	-	7470A	PPB	14 Apr 2017 12:42:53	U	SMPL
2655-13LX5 - 1		348	0.0840	-	7470A	PPB	14 Apr 2017 12:45:00	U	SMPL
2655-13A - 1		17506	4.8018	-	7470A	PPB	14 Apr 2017 12:47:07	U	SMPL
2670-03LX5 - 1		-6	-0.0133	-	7470A	PPB	14 Apr 2017 12:49:14	U	SMPL
2670-03A - 1		14601	4.0030	-	7470A	PPB	14 Apr 2017 12:51:21	U	SMPL
CCV21 - 1	CCV21	18368	5.0388	-	7470A	PPB	14 Apr 2017 12:53:28	U	SMPL
CCB21 - 1	CCB21	-39	-0.0224	-	7470A	PPB	14 Apr 2017 12:55:35	U	SMPL