



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86894

Temp.: 16.3 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/14/2017	12:05
2	CAL2	1	Water			7.01	04/14/2017	12:06
3	CAL3	1	Water			10.03	04/14/2017	12:07
4	ICV	1	Water			7.00	04/14/2017	12:08
5	CCV1	1	Water			2.02	04/14/2017	12:10
6	I2698-01	1	Water			4.23	04/14/2017	12:11
7	I2698-01DUP	1	Water			4.24	04/14/2017	12:12
8	CCV2	1	Water			12.01	04/14/2017	12:13

2386894

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I2698

WorkList ID : 97444

Date : 4/14/2017 11:48:14 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2698-01	pH	Cool 4 deg C	STER12	P12	CELLPOOLWATER	04/13/2017	SM 4500-PH B

Date/Time

04-14-17 11:48 AM

Received by:

JP

Relinquished by:

CS

Date/Time

04-14-17

11:00 AM

Received by:

JP

Relinquished by:

JP