


7470A

L386896
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36211



A= 0.0000e+000

B= 2.7573e-004

C= -1.0129e-002

Rho= 0.9999688

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	21	0.000	21				
0.2	0.200	0.175	-0.025	671	0.0 %	671				
2.5	2.500	2.544	0.044	9262	0.0 %	9262				
5.0	5.000	4.975	-0.025	18081	0.0 %	18081				
7.5	7.500	7.536	0.036	27367	0.0 %	27367				
10.0	10.000	9.974	-0.026	36211	0.0 %	36211				

LB86896 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Apr 2017 13:49:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.0000	7470A	PPB	14 Apr 2017 13:51:48	S	Std
0.2 - 1	50.2	671	-	0.2000	7470A	PPB	14 Apr 2017 13:53:56	S	Std
2.5 - 1	52.5	9262	-	2.5000	7470A	PPB	14 Apr 2017 13:56:04	S	Std
5.0 - 1	55	18081	-	5.0000	7470A	PPB	14 Apr 2017 13:58:12	S	Std
7.5 - 1	57.5	27367	-	7.5000	7470A	PPB	14 Apr 2017 14:00:20	S	Std
10.0 - 1	510	36211	-	10.0000	7470A	PPB	14 Apr 2017 14:02:27	S	Std
ICV20 - 1	ICV20	14252	3.9196	-	7470A	PPB	14 Apr 2017 14:05:04	U	SMPL
ICB20 - 1	ICB20	-72	-0.0300	-	7470A	PPB	14 Apr 2017 14:07:10	U	SMPL
CCV22 - 1	CCV22	18119	4.9859	-	7470A	PPB	14 Apr 2017 14:09:17	U	SMPL
CCB22 - 1	CCB22	-70	-0.0294	-	7470A	PPB	14 Apr 2017 14:11:23	U	SMPL
CRA - 1	CRA	633	0.1644	-	7470A	PPB	14 Apr 2017 14:13:29	U	SMPL
HighStd - 1	HighStd	36322	10.0051	-	7470A	PPB	14 Apr 2017 14:15:36	U	SMPL
ChkStd - 1	ChkStd	25423	6.9998	-	7470A	PPB	14 Apr 2017 14:17:43	U	SMPL
PB98262BL - 1	PBW	-107	-0.0396	-	7470A	PPB	14 Apr 2017 14:19:51	U	SMPL
PB98262BS - 1	LCSW	13844	3.8071	-	7470A	PPB	14 Apr 2017 14:22:00	U	SMPL
PB98262TB - 1	PB98247TB	-58	-0.0261	-	7470A	PPB	14 Apr 2017 14:24:08	U	SMPL
I2533-02 - 1	CT-MIXCOMP-1	6	-0.0085	-	7470A	PPB	14 Apr 2017 14:26:16	U	SMPL
I2533-02DUP - 1	CT-MIXCOMP-1DUP	13	-0.0065	-	7470A	PPB	14 Apr 2017 14:28:24	U	SMPL
I2533-02MS - 1	CT-MIXCOMP-1MS	13668	3.7586	-	7470A	PPB	14 Apr 2017 14:30:33	U	SMPL
I2533-02MSD - 1	CT-MIXCOMP-1MSD	14307	3.9348	-	7470A	PPB	14 Apr 2017 14:32:39	U	SMPL
CCV23 - 1	CCV23	18029	4.9611	-	7470A	PPB	14 Apr 2017 14:34:46	U	SMPL
CCB23 - 1	CCB23	-93	-0.0358	-	7470A	PPB	14 Apr 2017 14:36:53	U	SMPL
I2533-02LX5 - 1		-3	-0.0110	-	7470A	PPB	14 Apr 2017 14:39:00	U	SMPL
I2533-02A - 1		14128	3.8854	-	7470A	PPB	14 Apr 2017 14:41:07	U	SMPL
CCV24 - 1	CCV24	18175	5.0013	-	7470A	PPB	14 Apr 2017 14:43:14	U	SMPL
CCB24 - 1	CCB24	-89	-0.0347	-	7470A	PPB	14 Apr 2017 14:45:21	U	SMPL