



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86903

Temp.: 19.3 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/14/2017	17:35
2	CAL2	1	Water			7.02	04/14/2017	17:36
3	CAL3	1	Water			10.03	04/14/2017	17:37
4	ICV	1	Water			6.99	04/14/2017	17:38
5	CCV1	1	Water			2.01	04/14/2017	17:39
6	I2724-01	1	Water			12.52	04/14/2017	17:40
7	I2724-01DUP	1	Water			12.53	04/14/2017	17:41
8	CCV2	1	Water			12.02	04/14/2017	17:42

1386903

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I2724

WorkList ID : 97458

Date : 4/14/2017 5:25:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2724-01-C	pH	Cool 4 deg C	LIR001	P33	TWA-EFFE-GW	04/14/2017	SM 4500-PH B

Date/Time 04-14-17 6:25 PM
 Received by: JR
 Relinquished by: JR

Date/Time 04-14-17 6:30 PM
 Received by: JR
 Relinquished by: JR