



Analytical Summary Report

Analysis Method:	SM4500-O G	Reviewed By:	Indiana
Parameter:	Dissolved Oxygen	Supervisor Review By:	apatel
Run Number:	LB86905	MANGANOUS SULFATE SOLUTION:	W2191
BOD Water:	WP54161	Alkaline Iodide Azide:	W2035
Starch:	W2121	Sodium Thiosulfate, 0.025N:	W2169
Conc Sulfuric Acid:	W2151	pH Strips:	W2106
		Zero DO Std:	WP54227

Seq	LabID	ClientID	Sample Vol (ml)	Initial Reading (mL)	Final Reading (mL)	Diff (mg/L)	Result (mg/L)
1	WINKLER1	WINKLER1	300	0.00	8.50	8.50	8.65
2	WINKLER2	WINKLER2	300	8.50	17.30	8.80	

Calibrate DO meter with average reading : 8.65

Zero DO Reading : 0.13 Calibrated to : 0.00

Seq	LabID	ClientID	DL	Sample Vol (ml)	Temperature (°C)	Result (mg/L)	Anal Date	Anal Time
1	LB86905BL	LB86905BL	1	300	20.0	0.00	04/14/2017	17:30
2	I2710-01	SW16A	1	300	19.1	13.58	04/14/2017	17:32
3	I2710-02	SW13	1	300	19.2	12.96	04/14/2017	17:34
4	I2710-03	SW12	1	300	18.5	11.92	04/14/2017	17:36
5	I2710-04	SW3	1	300	18.9	12.98	04/14/2017	17:38
6	I2710-07	SW11	1	300	18.9	13.16	04/14/2017	17:40
7	I2710-08	SW6	1	300	19.2	12.58	04/14/2017	17:42
8	I2710-09	SW7	1	300	18.7	12.30	04/14/2017	17:44
9	I2710-10	SW5	1	300	19.0	11.85	04/14/2017	17:46
10	I2710-11	SW9	1	300	19.2	13.43	04/14/2017	17:48
11	I2710-12	SW4	1	300	18.8	13.45	04/14/2017	17:50
12	I2710-13	SW17	1	300	18.7	13.74	04/14/2017	17:52
13	I2710-14	SW14	1	300	18.8	13.16	04/14/2017	17:54
14	I2710-14DUP	SW14DUP	1	300	18.8	13.17	04/14/2017	17:56

LB 86905

WORKLIST(Hardcopy Internal Chain)

WorkList Name : DO011417

WorkList ID : 97450

Date : 4/14/2017 3:42:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/15/2017	Water	I2710-01 - B	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW16A	04/14/2017	SM4500-O G
04/15/2017	Water	I2710-02 - C	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW13	04/14/2017	SM4500-O G
04/15/2017	Water	I2710-03 - D	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW12	04/14/2017	SM4500-O G
04/15/2017	Water	I2710-04 - B	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW3	04/14/2017	SM4500-O G
04/15/2017	Water	I2710-07 - B	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW11	04/14/2017	SM4500-O G
04/15/2017	Water	I2710-08 - D	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW6	04/14/2017	SM4500-O G
04/15/2017	Water	I2710-09 - C	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW7	04/14/2017	SM4500-O G
04/15/2017	Water	I2710-10 - B	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW5	04/14/2017	SM4500-O G
04/15/2017	Water	I2710-11 - C	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW9	04/14/2017	SM4500-O G
04/15/2017	Water	I2710-12 - C	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW4	04/14/2017	SM4500-O G
04/15/2017	Water	I2710-13 - B	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW17	04/14/2017	SM4500-O G
04/15/2017	Water	I2710-14 - B	Dissolved Oxygen	Cool 4 deg C	TETR19	P51	SW14	04/14/2017	SM4500-O G

Date/Time 04/14/17 17:20

Received by: IB

Relinquished by: CB

Date/Time 04/14/17 18:05

Received by: IB

Relinquished by: IB