



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86906

Temp.: 23.2 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/12/2017	17:20
2	CAL2	1	Water			7.01	04/12/2017	17:21
3	CAL3	1	Water			10.03	04/12/2017	17:22
4	ICV	1	Water			7.01	04/12/2017	17:23
5	CCV1	1	Water			2.01	04/12/2017	17:40
6	I2671-04	1	Water			6.56	04/12/2017	17:41
7	I2671-04DUP	1	Water			6.57	04/12/2017	17:42
8	I2671-07	1	Water			5.01	04/12/2017	17:43
9	CCV2	1	Water			12.02	04/12/2017	17:44

28386906

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I2671 WorkList ID : 97459 Date : 4/12/2017 4:59:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/12/2017	Water	I2671-04	Corrosivity	Cool 4 deg C	PSEG01	O52	DRUMS-(2-7)	04/12/2017	9040C
04/12/2017	Water	I2671-07	Corrosivity	Cool 4 deg C	PSEG01	O52	M-15-M-18	04/12/2017	9040C

Date/Time 04-12-17 4:57 PM
Received by:
Relinquished by:

04-12-17 4:59 PM
Received by:
Relinquished by: