



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86907

Temp.: 24.4 °C

BalanceID: WC SC-4

SLOPE.: 98.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/14/2017	17:00
2	CAL2	1	Water			7.02	04/14/2017	17:01
3	CAL3	1	Water			10.03	04/14/2017	17:02
4	ICV	1	Water			7.01	04/14/2017	17:03
5	CCV1	1	Water			2.02	04/14/2017	17:05
6	I2718-02	1	Solid	20.03	20	8.16	04/14/2017	17:06
7	I2718-02DUP	1	Solid	20.05	20	8.17	04/14/2017	17:08
8	I2718-04	1	Solid	20.03	20	7.88	04/14/2017	17:10
9	CCV2	1	Water			12.02	04/14/2017	17:11

2086907

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2718 WorkList ID : 97454 Date : 4/14/2017 4:24:22 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2718-02	Corrosivity	Cool 4 deg C	ENTA04	P21	WC-B48-B49-001	04/13/2017	9045C
	Solid	I2718-04	Corrosivity	Cool 4 deg C	ENTA04	P21	WC-B60-B61-001	04/14/2017	9045C

Date/Time 04-14-17 11:24 AM
Received by: JS
Relinquished by: JS

Date/Time 04-14-17 7:00 PM
Received by: JS
Relinquished by: JS