



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB86909

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP54223
Turbidity std, 40NTU	WP54217
turbidity std 5 NTU	WP54220
turbidity std 20 NTU	WP54218
Turbidity std 1 NTU	WP54221
turbidity std 0 NTU	WP54216
turbidity std 10 NTU	WP54219
10 NTU Standard 500 ml	W2152

Intercept: 0.0733

Slope: 0.9701

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.080	04/14/2017	17:30
2	CAL2	1	1	1.049	04/14/2017	17:32
3	CAL3	5	1	4.899	04/14/2017	17:35
4	CAL4	10	1	9.850	04/14/2017	17:37
5	CAL5	20	1	19.400	04/14/2017	17:40
6	CAL6	40	1	38.900	04/14/2017	17:42



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB86909

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.700	9.923	04/14/2017	17:44
2	ICB		1	0.075	0.002	04/14/2017	17:47
3	CCV1	10	1	9.970	10.202	04/14/2017	17:50
4	CCB1		1	0.072	-0.001	04/14/2017	17:52
5	LB86909BL		1	0.081	0.008	04/14/2017	17:54
6	I2710-01		1	2.240	2.233	04/14/2017	17:57
7	I2710-01DUP		1	2.230	2.223	04/14/2017	18:00
8	I2710-02		1	5.700	5.800	04/14/2017	18:02
9	I2710-03		1	5.750	5.852	04/14/2017	18:05
10	I2710-04		1	2.740	2.749	04/14/2017	18:07
11	I2710-07		1	4.200	4.254	04/14/2017	18:09
12	I2710-08		1	2.360	2.357	04/14/2017	18:12
13	I2710-09		1	20.300	20.850	04/14/2017	18:15
14	I2710-10		1	1.330	1.295	04/14/2017	18:17
15	CCV2	10	1	9.980	10.212	04/14/2017	18:20
16	CCB2		1	0.074	0.001	04/14/2017	18:22
17	I2710-11		1	6.130	6.243	04/14/2017	18:25
18	I2710-12		1	2.740	2.749	04/14/2017	18:27
19	I2710-13		1	1.900	1.883	04/14/2017	18:30
20	I2710-14		1	4.620	4.687	04/14/2017	18:32
21	CCV3	10	1	9.850	10.078	04/14/2017	18:35
22	CCB3		1	0.074	0.001	04/14/2017	18:37

LB 86909

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TUR041417

WorkList ID : 97452

Date : 4/14/2017 4:02:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2710-01	A Turbidity	Cool 4 deg C	TETR19	P51	SW16A	04/14/2017	SM2130 B
	Water	I2710-02	D Turbidity	Cool 4 deg C	TETR19	P51	SW13	04/14/2017	SM2130 B
	Water	I2710-03	C Turbidity	Cool 4 deg C	TETR19	P51	SW12	04/14/2017	SM2130 B
	Water	I2710-04	A Turbidity	Cool 4 deg C	TETR19	P51	SW3	04/14/2017	SM2130 B
	Water	I2710-07	B Turbidity	Cool 4 deg C	TETR19	P51	SW11	04/14/2017	SM2130 B
	Water	I2710-08	C Turbidity	Cool 4 deg C	TETR19	P51	SW6	04/14/2017	SM2130 B
	Water	I2710-09	D Turbidity	Cool 4 deg C	TETR19	P51	SW7	04/14/2017	SM2130 B
	Water	I2710-10	A Turbidity	Cool 4 deg C	TETR19	P51	SW5	04/14/2017	SM2130 B
	Water	I2710-11	D Turbidity	Cool 4 deg C	TETR19	P51	SW9	04/14/2017	SM2130 B
	Water	I2710-12	B Turbidity	Cool 4 deg C	TETR19	P51	SW4	04/14/2017	SM2130 B
	Water	I2710-13	A Turbidity	Cool 4 deg C	TETR19	P51	SW17	04/14/2017	SM2130 B
	Water	I2710-14	A Turbidity	Cool 4 deg C	TETR19	P51	SW14	04/14/2017	SM2130 B

Date/Time 4/14/17 5:15 PM

Received by: AK

Relinquished by: AK

Date/Time 4/14/17 19:10

Received by: AK

Relinquished by: AK