

AK
4/18/17

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
CCV	9.0509			TOC	
CCB	0.1195			TOC	
LB86917BLW	0.0427	0.0393	92.07	TOC	
LB86917BSW.....	9.3404...	0.1247..	1.33...	TOC	..
I2699-02	-0.6635	1.3991	-210.86	TOC	
I2699-02DUP	-0.8393	1.6012	-190.77	TOC	
I2699-01DLX50.....	42.8075...	0.6453..	1.51...	TOC	..
I2699-03DLX50	18.1225	0.0215	0.12	TOC	
I2699-04DLX200	11.0363	0.0173	0.16	TOC	
I2699-01DLX200.....	11.4671...	0.0114..	0.10...	TOC	..
I2699-02MS	9.8696	0.0502	0.51	TOC	
I2699-02MSD	9.3916	0.3780	4.03	TOC	
CCV.....	9.1674...	..	...	TOC	..
CCB	0.1518			TOC	

Method ID	Sample Type	Vial	Timestamp	Message
=====	=====	=====	=====	=====
TOC 0 - 20 ppmC	Sample	2	2017/04/17 10:45	
TOC 0 - 20 ppmC	Sample	3	2017/04/17 10:58	Low Sample Detected
TOC 0 - 20 ppmC	Sample	4	2017/04/17 11:50	Low Sample Detected
TOC 0 - 20 ppmC	...Sample	.. 5..	2017/04/17 12:17	..
TOC 0 - 20 ppmC	Sample	6	2017/04/17 13:08	Low Sample Detected
TOC 0 - 20 ppmC	Sample	7	2017/04/17 13:24	Low Sample Detected
TOC 0 - 20 ppmC	...Sample	.. 10..	2017/04/17 13:46	..
TOC 0 - 20 ppmC	Sample	11	2017/04/17 14:06	
TOC 0 - 20 ppmC	Sample	14	2017/04/17 14:41	
TOC 0 - 20 ppmC	...Sample	.. 13..	2017/04/17 15:01	..
TOC 0 - 20 ppmC	Sample	8	2017/04/17 16:31	
TOC 0 - 20 ppmC	Sample	9	2017/04/17 16:50	
TOC 0 - 20 ppmC	...Sample	.. 2..	2017/04/17 17:34	..
TOC 0 - 20 ppmC	Sample	3	2017/04/17 17:47	Low Sample Detected

Cal. Curve ID: TOC WATER 0-20PPM
Created: 2017/03/20 12:28
Calibration Factor (m): 1.027e+05
Y Intercept (b): 14049
r-squared: 0.99964

Standard ID	Y Raw Data	X Expected ug C	Measured ug C	Message	Date & Time
0.0PPM	8869	0.000	-0.050		2017/03/20 11:05
0.5PPM	32419	0.250	0.179		2017/03/20 11:19
1.0PPM	63052	0.500	0.477		2017/03/20 11:32
2.0PPM	120171	1.000	1.033		2017/03/20 11:45
5.0PPM	280664	2.500	2.596		2017/03/20 11:59
10.0PPM	535837	5.000	5.080		2017/03/20 12:13
20.0PPM	1034598	10.000	9.936		2017/03/20 12:27

3-22-17
AK

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
0.0PPM	8869			TOC	
0.5PPM	32419			TOC	
1.0PPM	63052			TOC	
2.0PPM.....	120171...	..	...	TOC	..
5.0PPM	280664			TOC	
10.0PPM	535837			TOC	
20.0PPM.....	1034598...	..	...	TOC	..
ICV	10.6158			TOC	
ICB	0.2279			TOC	

3-22-17
AK

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	TOC Standard	1	2017/03/20 11:05	
TOC 0 - 20 ppmC	TOC Standard	2	2017/03/20 11:19	
TOC 0 - 20 ppmC	TOC Standard	3	2017/03/20 11:32	
TOC 0 - 20 ppmC	...TOC Standard	.. 4..	2017/03/20 11:45	..
TOC 0 - 20 ppmC	TOC Standard	5	2017/03/20 11:59	
TOC 0 - 20 ppmC	TOC Standard	6	2017/03/20 12:13	
TOC 0 - 20 ppmC	...TOC Standard	.. 7..	2017/03/20 12:27	..
TOC 0 - 20 ppmC	Sample	8	2017/03/20 12:42	
TOC 0 - 20 ppmC	Sample	9	2017/03/20 12:54	