



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86921

Temp.: 22.3 °C

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/17/2017	13:30
2	CAL2	1	Water			7.01	04/17/2017	13:31
3	CAL3	1	Water			10.02	04/17/2017	13:32
4	ICV	1	Water			6.99	04/17/2017	13:33
5	CCV1	1	Water			2.02	04/17/2017	13:35
6	I2730-03	1	Water			4.43	04/17/2017	13:36
7	I2730-03DUP	1	Water			4.44	04/17/2017	13:37
8	I2731-03	1	Water			4.52	04/17/2017	13:38
9	CCV2	1	Water			12.01	04/17/2017	13:40

LB 86921

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I2730

WorkList ID : 97495

Date : 4/17/2017 1:14:32 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2730-03	pH	Cool 4 deg C	STER16		EXCELL	04/12/2017	SM 4500-PH B
	Water	I2731-03	pH	Cool 4 deg C	STER16		MAINCELL	04/11/2017	SM 4500-PH B

Date/Time 04-17-17 1:14 PM
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 04-17-17 2:05 PM
 Received by: [Signature]
 Relinquished by: [Signature]