



### Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86924

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/17/2017	14:30
2	CAL2	1	Water			7.00	04/17/2017	14:31
3	CAL3	1	Water			10.04	04/17/2017	14:32
4	ICV	1	Water			7.01	04/17/2017	14:33
5	CCV1	1	Water			2.02	04/17/2017	14:35
6	I2732-01	1	Solid	20.03	20	7.26	04/17/2017	14:36
7	I2732-01DUP	1	Solid	20.04	20	7.27	04/17/2017	14:37
8	I2735-01	1	Solid	20.05	20	7.14	04/17/2017	14:38
9	CCV2	1	Water			12.02	04/17/2017	14:40

28869224

### WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i2732

WorkList ID : 97499

Date : 4/17/2017 2:14:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2732-01	pH	Cool 4 deg C	TULL02	P51	EME-COMPOST-1	04/17/2017	9045C
	Solid	I2735-01	pH	Cool 4 deg C	TULL02	P51	EME-COMPOST-1	04/17/2017	9045C

Date/Time 04-17-2017 PM  
 Received by: JP  
 Relinquished by: JK

Date/Time 04-17-17 3:40PM  
 Received by: JK  
 Relinquished by: JP