



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: JIGNESHDate: 04/17/2017DATE & TIME IN: 04/17/2017 15:35OVEN TEMP: 104 °CRun Number: LB86925DATE & TIME OUT: 04/17/2017 18:35OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB86925BL	LB86925BL	1.3287	1.3287	100	1.3287	1.3287	1.3287	0	0
2	LB86925BS	LB86925BS	1.4762	1.4762	100	1.5308	1.5308	1.5308	0.0546	546
3	I2738-01	NORTHPORTALWALL-HYDRODEMO	1.4196	1.4196	150	1.7228	1.7228	1.7228	0.3032	2021.3
4	I2738-01DUP	NORTHPORTALWALL-HYDRODEMODU	1.4186	1.4186	150	1.7371	1.7371	1.7371	0.3185	2123.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

2B86925

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss i2738

WorkList ID : 97502

Date : 4/17/2017 3:18:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2738-01	TSS	Cool 4 deg C	TULL01	P12	NORTHPORTALWALL-HYDR	03/27/2017	SM2540 D

Date/Time 04-17-17 3:18 PM
 Received by:
 Relinquished by:

Date/Time 04-17-17 4:30 PM
 Received by:
 Relinquished by: