



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86932

Temp.: 23.1 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/17/2017	17:40
2	CAL2	1	Water			7.02	04/17/2017	17:41
3	CAL3	1	Water			10.02	04/17/2017	17:42
4	ICV	1	Water			7.01	04/17/2017	17:43
5	CCV1	1	Water			2.02	04/17/2017	17:45
6	I2742-01	1	Water			4.61	04/17/2017	17:46
7	I2742-01DUP	1	Water			4.62	04/17/2017	17:47
8	CCV2	1	Water			12.02	04/17/2017	17:48

LB86932

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 12742

WorkList ID : 97504

Date : 4/17/2017 5:29:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2742-01	pH	Cool 4 deg C	STER06	P51	POOLWATER	04/07/2017	SM 4500-PH B

Date/Time 04-17-17 5:29 PM
Received by: JB
Relinquished by:

Date/Time 04-17-17 6:20 PM
Received by: CP
Relinquished by: JP