



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB86943

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP53540
calibration std. phosphate 1 ppm	WP54266
calibration std. phosphate 0.5 ppm	WP54265
calibration std. phosphate 0.3 ppm	WP54264
calibration std. 0 ppm	WP54261
phosphate ICV std.	WP54268
calibration std. phosphate 0.1 ppm	WP54263
calibration std. phosphate 0.05 ppm	WP54262
Combined reagent	WP54270
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
phosphate CCV std.	WP54267

Intercept: -0.0005

Slope: 0.6487

Regression: 0.9998

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.001	04/18/2017	12:01
2	CAL2	0.05	1	50	50	0.032	04/18/2017	12:01
3	CAL3	0.10	1	50	50	0.065	04/18/2017	12:02
4	CAL4	0.30	1	50	50	0.186	04/18/2017	12:02
5	CAL5	0.50	1	50	50	0.331	04/18/2017	12:03
6	CAL6	1.00	1	50	50	0.647	04/18/2017	12:03



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB86943

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.330	0.509	04/18/2017	12:04
2	ICB		1	50	50	0.000	0.001	04/18/2017	12:04
3	CCV1	0.50	1	50	50	0.331	0.511	04/18/2017	12:05
4	CCB1		1	50	50	0.000	0.001	04/18/2017	12:05
5	PB98312BL		1	50	50	0.000	0.001	04/18/2017	12:06
6	PB98312BS	0.50	1	50	50	0.335	0.517	04/18/2017	12:06
7	I2710-01		1	50	50	0.040	0.062	04/18/2017	12:07
8	I2710-01DUP		1	50	50	0.041	0.064	04/18/2017	12:07
9	I2710-01MS	0.50	1	50	50	0.348	0.537	04/18/2017	12:08
10	I2710-01MSD	0.50	1	50	50	0.345	0.533	04/18/2017	12:08
11	I2710-02		1	50	50	0.049	0.076	04/18/2017	12:09
12	I2710-03		1	50	50	0.024	0.038	04/18/2017	12:09
13	I2710-04		1	50	50	0.047	0.073	04/18/2017	12:10
14	I2710-07		1	50	50	0.049	0.076	04/18/2017	12:10
15	CCV2	0.50	1	50	50	0.329	0.508	04/18/2017	12:11
16	CCB2		1	50	50	0.000	0.001	04/18/2017	12:11
17	I2710-08		1	50	50	0.038	0.059	04/18/2017	12:12
18	I2710-09		1	50	50	0.086	0.133	04/18/2017	12:12
19	I2710-10		1	50	50	0.032	0.050	04/18/2017	12:13
20	I2710-11		1	50	50	0.043	0.067	04/18/2017	12:13
21	I2710-12		1	50	50	0.085	0.132	04/18/2017	12:14
22	I2710-13		1	50	50	0.074	0.115	04/18/2017	12:14
23	I2710-14		1	50	50	0.052	0.081	04/18/2017	12:15
24	CCV3	0.50	1	50	50	0.330	0.509	04/18/2017	12:15
25	CCB3		1	50	50	0.000	0.001	04/18/2017	12:16