

245.1


LB 86946

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

35032



A= 0.0000e+000

B= 2.8813e-004

C= -5.7737e-003

Rho= 0.9998063

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	29	0.000	29				
0.2	0.200	0.216	0.016	769	0.0 %	769				
2.5	2.500	2.504	0.004	8712	0.0 %	8712				
5.0	5.000	5.038	0.038	17505	0.0 %	17505				
7.5	7.500	7.351	-0.149	25533	0.0 %	25533				
10.0	10.000	10.088	0.088	35032	0.0 %	35032				

LB86946 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 18 Apr 2017 15:23:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	29	-	0.0000	245.1	PPB	18 Apr 2017 15:30:48	S	Std
0.2 - 1	50.2	769	-	0.2000	245.1	PPB	18 Apr 2017 15:32:56	S	Std
2.5 - 1	52.5	8712	-	2.5000	245.1	PPB	18 Apr 2017 15:35:04	S	Std
5.0 - 1	55	17505	-	5.0000	245.1	PPB	18 Apr 2017 15:37:13	S	Std
7.5 - 1	57.5	25533	-	7.5000	245.1	PPB	18 Apr 2017 15:39:20	S	Std
10.0 - 1	510	35032	-	10.0000	245.1	PPB	18 Apr 2017 15:41:27	S	Std
ICV23 - 1	ICV23	13945	4.0122	-245.1		PPB	18 Apr 2017 15:44:02	U	SMPL
ICB23 - 1	ICB23	-6	-0.0075	-245.1		PPB	18 Apr 2017 15:46:08	U	SMPL
CCV40 - 1	CCV40	17456	5.0239	-245.1		PPB	18 Apr 2017 15:48:15	U	SMPL
CCB40 - 1	CCB40	-16	-0.0104	-245.1		PPB	18 Apr 2017 15:50:21	U	SMPL
CRA - 1	CRA	762	0.2138	-245.1		PPB	18 Apr 2017 15:52:28	U	SMPL
HighStd - 1	HighStd	34736	10.0028	-245.1		PPB	18 Apr 2017 15:54:34	U	SMPL
ChkStd - 1	ChkStd	24226	6.9745	-245.1		PPB	18 Apr 2017 15:56:41	U	SMPL
PB98316BL - 1	PBW	-33	-0.0153	-245.1		PPB	18 Apr 2017 15:58:49	U	SMPL
PB98316BS - 1	LCSW	14041	4.0399	-245.1		PPB	18 Apr 2017 16:00:57	U	SMPL
I2562-03 - 1	0206	8703	2.5018	-245.1		PPB	18 Apr 2017 16:03:04	U	SMPL
CCV41 - 1	CCV41	17289	4.9757	-245.1		PPB	18 Apr 2017 16:11:36	U	SMPL
CCB41 - 1	CCB41	-16	-0.0104	-245.1		PPB	18 Apr 2017 16:13:43	U	SMPL