



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86947

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/18/2017	16:30
2	CAL2	1	Water			7.01	04/18/2017	16:31
3	CAL3	1	Water			10.03	04/18/2017	16:32
4	ICV	1	Water			7.01	04/18/2017	16:33
5	CCV1	1	Water			2.01	04/18/2017	16:35
6	I2751-03	1	Solid	20.03	20	6.34	04/18/2017	16:36
7	I2751-03DUP	1	Solid	20.04	20	6.35	04/18/2017	16:37
8	I2751-06	1	Solid	20.02	20	7.01	04/18/2017	16:38
9	I2751-09	1	Solid	20.05	20	7.36	04/18/2017	16:40
10	I2752-03	1	Solid	20.03	20	7.69	04/18/2017	16:41
11	I2752-06	1	Solid	20.04	20	7.46	04/18/2017	16:42
12	CCV2	1	Water			12.02	04/18/2017	16:43

LB 86947

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2751 WorkList ID : 97542 Date : 4/18/2017 2:34:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/21/2017	Solid	I2751-03	Corrosivity	Cool 4 deg C	PSEG01	P22	15-40-COMP	04/18/2017	9045C
04/21/2017	Solid	I2751-06	Corrosivity	Cool 4 deg C	PSEG01	P22	RT-1494-COMP	04/18/2017	9045C
04/21/2017	Solid	I2751-09	Corrosivity	Cool 4 deg C	PSEG01	P22	VNJ-209-COMP	04/18/2017	9045C
04/21/2017	Solid	I2752-03	Corrosivity	Cool 4 deg C	PSEG01	P21	15-50-COMP	04/18/2017	9045C
04/21/2017	Solid	I2752-06	Corrosivity	Cool 4 deg C	PSEG01	P21	VNJ-215-COMP	04/18/2017	9045C

Date/Time 04-18-17 2:34 PM
 Received by: SR
 Relinquished by: SR

Date/Time 04-18-17 5:30 PM
 Received by: SR
 Relinquished by: SR