

7470A


L386951

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37846



A= 0.0000e+000

B= 2.6547e-004

C= -1.0408e-002

Rho= 0.9999338

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	194	0.000	194				
0.2	0.200	0.213	0.013	841	0.0 %	841				
2.5	2.500	2.482	-0.018	9390	0.0 %	9390				
5.0	5.000	4.915	-0.085	18555	0.0 %	18555				
7.5	7.500	7.512	0.012	28337	0.0 %	28337				
10.0	10.000	10.036	0.036	37846	0.0 %	37846				

LB86951 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Apr 2017 16:30:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	194	-	0.00007470A	PPB		18 Apr 2017 16:32:16	S	Std
0.2 - 1	50.2	841	-	0.20007470A	PPB		18 Apr 2017 16:34:24	S	Std
2.5 - 1	52.5	9390	-	2.50007470A	PPB		18 Apr 2017 16:36:33	S	Std
5.0 - 1	55	18555	-	5.00007470A	PPB		18 Apr 2017 16:38:41	S	Std
7.5 - 1	57.5	28337	-	7.50007470A	PPB		18 Apr 2017 16:40:49	S	Std
10.0 - 1	510	37846	-	10.00007470A	PPB		18 Apr 2017 16:42:56	S	Std
ICV24 - 1	ICV24	14577	3.8593	-7470A	PPB		18 Apr 2017 16:45:29	U	SMPL
ICB24 - 1	ICB24	132	0.0246	-7470A	PPB		18 Apr 2017 16:47:35	U	SMPL
CCV42 - 1	CCV42	18616	4.9315	-7470A	PPB		18 Apr 2017 16:49:42	U	SMPL
CCB42 - 1	CCB42	122	0.0220	-7470A	PPB		18 Apr 2017 16:51:48	U	SMPL
CRA - 1	CRA	845	0.2139	-7470A	PPB		18 Apr 2017 16:53:55	U	SMPL
HighStd - 1	HighStd	37717	10.0021	-7470A	PPB		18 Apr 2017 16:56:02	U	SMPL
ChkStd - 1	ChkStd	26048	6.9044	-7470A	PPB		18 Apr 2017 16:58:09	U	SMPL
PB98287BL - 1	PBW	29	-0.0027	-7470A	PPB		18 Apr 2017 17:00:17	U	SMPL
PB98287BS - 1	LCSW	14305	3.7871	-7470A	PPB		18 Apr 2017 17:02:24	U	SMPL
PB98287TB - 1	PB98265TB	39	-0.0001	-7470A	PPB		18 Apr 2017 17:04:32	U	SMPL
I2701-02 - 1	S8-9004	101	0.0164	-7470A	PPB		18 Apr 2017 17:06:40	U	SMPL
I2718-02 - 1	WC-B48-B49-001	114	0.0199	-7470A	PPB		18 Apr 2017 17:08:48	U	SMPL
I2718-02DUP - 1	WC-B48-B49-001DUP	127	0.0233	-7470A	PPB		18 Apr 2017 17:10:56	U	SMPL
I2718-02MS - 1	WC-B48-B49-001MS	13485	3.5694	-7470A	PPB		18 Apr 2017 17:13:03	U	SMPL
CCV43 - 1	CCV43	18179	4.8155	-7470A	PPB		18 Apr 2017 17:15:09	U	SMPL
CCB43 - 1	CCB43	123	0.0222	-7470A	PPB		18 Apr 2017 17:17:16	U	SMPL
I2718-02MSD - 1	WC-B48-B49-001MSD	13492	3.5712	-7470A	PPB		18 Apr 2017 17:19:23	U	SMPL
I2718-04 - 1	WC-B60-B61-001	345	0.0812	-7470A	PPB		18 Apr 2017 17:21:30	U	SMPL
I2718-02LX5 - 1		143	0.0276	-7470A	PPB		18 Apr 2017 17:23:37	U	SMPL
I2718-02A - 1		14598	3.8649	-7470A	PPB		18 Apr 2017 17:25:44	U	SMPL
CCV44 - 1	CCV44	18561	4.9169	-7470A	PPB		18 Apr 2017 17:27:51	U	SMPL
CCB44 - 1	CCB44	95	0.0148	-7470A	PPB		18 Apr 2017 17:29:59	U	SMPL