


7470A

LB 86958
INSTRUMENT ID: CV4

Linear

 μ Abs.:

35461



A= 0.0000e+000

B= 2.7893e-004

C= 4.7899e-002

Rho= 0.9995329

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.034	0.034	-50	0.000	-50				
0.2	0.200	0.250	0.050	724	0.0 %	724				
2.5	2.500	2.287	-0.213	8028	0.0 %	8028				
5.0	5.000	5.148	0.148	18284	0.0 %	18284				
7.5	7.500	7.542	0.042	26868	0.0 %	26868				
10.0	10.000	9.939	-0.061	35461	0.0 %	35461				

LB86958 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Apr 2017 10:19:04

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-50	-	0.0000	7470A	PPB	19 Apr 2017 10:41:32	S	Std
0.2 - 1	50.2	724	-	0.2000	7470A	PPB	19 Apr 2017 10:43:40	S	Std
2.5 - 1	52.5	8028	-	2.5000	7470A	PPB	19 Apr 2017 10:45:49	S	Std
5.0 - 1	55	18284	-	5.0000	7470A	PPB	19 Apr 2017 10:47:57	S	Std
7.5 - 1	57.5	26868	-	7.5000	7470A	PPB	19 Apr 2017 10:52:45	S	Std
10.0 - 1	110	35461	-	10.0000	7470A	PPB	19 Apr 2017 10:54:52	S	Std
ICV25 - 1	ICV25	14444	4.0768	-	7470A	PPB	19 Apr 2017 10:57:32	U	SMPL
ICB25 - 1	ICB25	-106	0.0183	-	7470A	PPB	19 Apr 2017 10:59:38	U	SMPL
CCV45 - 1	CCV45	17215	4.8497	-	7470A	PPB	19 Apr 2017 11:01:45	U	SMPL
CCB45 - 1	CCB45	-108	0.0178	-	7470A	PPB	19 Apr 2017 11:03:52	U	SMPL
CRA - 1	CRA	674	0.2359	-	7470A	PPB	19 Apr 2017 11:08:26	U	SMPL
HighStd - 1	HighStd	35947	10.0746	-	7470A	PPB	19 Apr 2017 11:10:32	U	SMPL
ChkStd - 1	ChkStd	25093	7.0471	-	7470A	PPB	19 Apr 2017 11:12:39	U	SMPL
PB98320BL - 1	PBW	-135	0.0102	-	7470A	PPB	19 Apr 2017 11:14:48	U	SMPL
PB98320BS - 1	LCSW	14037	3.9632	-	7470A	PPB	19 Apr 2017 11:16:56	U	SMPL
I2710-01 - 1	SW16A	-56	0.0323	-	7470A	PPB	19 Apr 2017 11:19:04	U	SMPL
I2710-02 - 1	SW13	26	0.0552	-	7470A	PPB	19 Apr 2017 11:21:12	U	SMPL
I2710-02DUP - 1	SW13DUP	73	0.0683	-	7470A	PPB	19 Apr 2017 11:25:51	U	SMPL
I2710-02MS - 1	SW13MS	15139	4.2706	-	7470A	PPB	19 Apr 2017 11:27:59	U	SMPL
I2710-02MSD - 1	SW13MSD	15001	4.2321	-	7470A	PPB	19 Apr 2017 11:30:05	U	SMPL
CCV46 - 1	CCV46	17356	4.8890	-	7470A	PPB	19 Apr 2017 11:32:12	U	SMPL
CCB46 - 1	CCB46	-82	0.0250	-	7470A	PPB	19 Apr 2017 11:34:19	U	SMPL
I2710-03 - 1	SW12	-4	0.0468	-	7470A	PPB	19 Apr 2017 11:36:26	U	SMPL
I2710-04 - 1	SW3	-7	0.0459	-	7470A	PPB	19 Apr 2017 11:38:33	U	SMPL
I2710-07 - 1	SW11	19	0.0532	-	7470A	PPB	19 Apr 2017 11:40:39	U	SMPL
I2710-08 - 1	SW6	13	0.0515	-	7470A	PPB	19 Apr 2017 11:42:46	U	SMPL
I2710-09 - 1	SW7	1677	0.5157	-	7470A	PPB	19 Apr 2017 11:44:54	U	SMPL
I2710-10 - 1	SW5	3	0.0487	-	7470A	PPB	19 Apr 2017 11:47:02	U	SMPL
I2710-11 - 1	SW9	20	0.0535	-	7470A	PPB	19 Apr 2017 11:49:10	U	SMPL
I2710-12 - 1	SW4	21	0.0538	-	7470A	PPB	19 Apr 2017 11:51:18	U	SMPL
I2710-13 - 1	SW17	66	0.0663	-	7470A	PPB	19 Apr 2017 11:53:27	U	SMPL
I2710-14 - 1	SW14	43	0.0599	-	7470A	PPB	19 Apr 2017 11:55:34	U	SMPL
CCV47 - 1	CCV47	16986	4.7858	-	7470A	PPB	19 Apr 2017 11:57:41	U	SMPL
CCB47 - 1	CCB47	-70	0.0284	-	7470A	PPB	19 Apr 2017 11:59:48	U	SMPL
PB98321BL - 1	PBW	-2	0.0473	-	7470A	PPB	19 Apr 2017 12:01:55	U	SMPL
PB98321BS - 1	LCSW	14145	3.9934	-	7470A	PPB	19 Apr 2017 12:04:03	U	SMPL
I2710-15 - 1	SW16A	-25	0.0409	-	7470A	PPB	19 Apr 2017 12:06:10	U	SMPL
I2710-16 - 1	SW13	0	0.0479	-	7470A	PPB	19 Apr 2017 12:08:17	U	SMPL
I2710-17 - 1	SW12	8	0.0501	-	7470A	PPB	19 Apr 2017 12:10:25	U	SMPL
I2710-18 - 1	SW3	16	0.0524	-	7470A	PPB	19 Apr 2017 12:12:32	U	SMPL
I2710-19 - 1	SW11	7	0.0499	-	7470A	PPB	19 Apr 2017 12:14:41	U	SMPL
I2710-20 - 1	SW6	16	0.0524	-	7470A	PPB	19 Apr 2017 12:16:49	U	SMPL
I2710-21 - 1	SW7	18	0.0529	-	7470A	PPB	19 Apr 2017 12:18:58	U	SMPL
I2710-22 - 1	SW5	17	0.0526	-	7470A	PPB	19 Apr 2017 12:21:05	U	SMPL
CCV48 - 1	CCV48	17364	4.8912	-	7470A	PPB	19 Apr 2017 12:23:12	U	SMPL
CCB48 - 1	CCB48	-86	0.0239	-	7470A	PPB	19 Apr 2017 12:25:19	U	SMPL
I2710-23 - 1	SW9	-1	0.0476	-	7470A	PPB	19 Apr 2017 12:27:26	U	SMPL
I2710-24 - 1	SW4	0	0.0479	-	7470A	PPB	19 Apr 2017 12:29:34	U	SMPL
I2710-25 - 1	SW17	48	0.0613	-	7470A	PPB	19 Apr 2017 12:31:41	U	SMPL
I2710-26 - 1	SW14	56	0.0635	-	7470A	PPB	19 Apr 2017 12:33:48	U	SMPL
I2719-01 - 1	VL-WWTP-001	43	0.0599	-	7470A	PPB	19 Apr 2017 12:35:56	U	SMPL
I2719-01DUP - 1	VL-WWTP-001DUP	27	0.0554	-	7470A	PPB	19 Apr 2017 12:38:03	U	SMPL
I2719-01MS - 1	VL-WWTP-001MS	14345	4.0491	-	7470A	PPB	19 Apr 2017 12:40:12	U	SMPL
I2719-01MSD - 1	VL-WWTP-001MSD	14820	4.1816	-	7470A	PPB	19 Apr 2017 12:42:20	U	SMPL
CCV49 - 1	CCV49	18260	5.1412	-	7470A	PPB	19 Apr 2017 12:44:28	U	SMPL
CCB49 - 1	CCB49	-100	0.0200	-	7470A	PPB	19 Apr 2017 12:46:35	U	SMPL
2710-02LX5 - 1		-10	0.0451	-	7470A	PPB	19 Apr 2017 12:48:42	U	SMPL
2710-02A - 1		14438	4.0751	-	7470A	PPB	19 Apr 2017 12:50:49	U	SMPL
2719-01LX5 - 1		-71	0.0281	-	7470A	PPB	19 Apr 2017 12:52:56	U	SMPL
2719-01A - 1		14642	4.1320	-	7470A	PPB	19 Apr 2017 12:55:03	U	SMPL
CCV50 - 1	CCV50	18056	5.0843	-	7470A	PPB	19 Apr 2017 12:57:11	U	SMPL
CCB50 - 1	CCB50	-112	0.0167	-	7470A	PPB	19 Apr 2017 12:59:18	U	SMPL