



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86963

Temp.: 18.4 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/19/2017	14:10
2	CAL2	1	Water			6.99	04/19/2017	14:11
3	CAL3	1	Water			10.02	04/19/2017	14:12
4	ICV	1	Water			7.01	04/19/2017	14:13
5	CCV1	1	Water			2.02	04/19/2017	14:15
6	I2763-01	1	Water			4.14	04/19/2017	14:16
7	I2763-01DUP	1	Water			4.15	04/19/2017	14:17
8	CCV2	1	Water			12.01	04/19/2017	14:18

188963

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i2763

WorkList ID : 97574

Date : 4/19/2017 1:58:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2763-01	pH	Cool 4 deg C	STER14	P23	POOLWATER	04/16/2016	SM 4500-PH B

Date/Time 04-19-17 2:00 PM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 04-19-17 3:10 PM
Received by: [Signature]
Relinquished by: [Signature]