

245.1


LB 86964

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34528



A= 0.0000e+000

B= 2.8951e-004

C= 1.3523e-002

Rho= 0.9999773

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	20	0.000	20				
0.2	0.200	0.218	0.018	708	0.0 %	708				
2.5	2.500	2.447	-0.053	8404	0.0 %	8404				
5.0	5.000	5.006	0.006	17245	0.0 %	17245				
7.5	7.500	7.500	-0.000	25858	0.0 %	25858				
10.0	10.000	10.010	0.010	34528	0.0 %	34528				

LB86964 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 19 Apr 2017 14:10:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	20	-	0.0000	245.1	PPB	19 Apr 2017 14:12:30	S	Std
0.2 - 1	50.2	708	-	0.2000	245.1	PPB	19 Apr 2017 14:14:38	S	Std
2.5 - 1	52.5	8404	-	2.5000	245.1	PPB	19 Apr 2017 14:16:47	S	Std
5.0 - 1	55	17245	-	5.0000	245.1	PPB	19 Apr 2017 14:18:55	S	Std
7.5 - 1	57.5	25858	-	7.5000	245.1	PPB	19 Apr 2017 14:21:03	S	Std
10.0 - 1	510	34528	-	10.0000	245.1	PPB	19 Apr 2017 14:25:36	S	Std
ICV26 - 1	ICV26	13287	3.8603	-245.1		PPB	19 Apr 2017 14:28:08	U	SMPL
ICB26 - 1	ICB26	-48	-0.0004	-245.1		PPB	19 Apr 2017 14:30:14	U	SMPL
CCV51 - 1	CCV51	17210	4.9960	-245.1		PPB	19 Apr 2017 14:32:21	U	SMPL
CCB51 - 1	CCB51	-46	0.0002	-245.1		PPB	19 Apr 2017 14:34:27	U	SMPL
CRA - 1	CRA	687	0.2124	-245.1		PPB	19 Apr 2017 14:36:33	U	SMPL
HighStd - 1	HighStd	34582	10.0254	-245.1		PPB	19 Apr 2017 14:38:40	U	SMPL
ChkStd - 1	ChkStd	23757	6.8914	-245.1		PPB	19 Apr 2017 14:40:47	U	SMPL
PB98345BL - 1	PBW	-98	-0.0148	-245.1		PPB	19 Apr 2017 14:42:55	U	SMPL
PB98345BS - 1	LCSW	13764	3.9984	-245.1		PPB	19 Apr 2017 14:45:03	U	SMPL
I2159-01 - 1	EFF-WW	-59	-0.0036	-245.1		PPB	19 Apr 2017 14:47:11	U	SMPL
I2159-01DUP - 1	EFF-WWDUP	7	0.0155	-245.1		PPB	19 Apr 2017 14:49:19	U	SMPL
I2159-01MS - 1	EFF-WWMS	13145	3.8191	-245.1		PPB	19 Apr 2017 14:53:59	U	SMPL
I2159-01MSD - 1	EFF-WWMSD	13444	3.9057	-245.1		PPB	19 Apr 2017 14:56:07	U	SMPL
CCV52 - 1	CCV52	17072	4.9561	-245.1		PPB	19 Apr 2017 14:58:14	U	SMPL
CCB52 - 1	CCB52	-46	0.0002	-245.1		PPB	19 Apr 2017 15:00:21	U	SMPL
I2159-01LX5 - 1		11	0.0167	-245.1		PPB	19 Apr 2017 15:02:27	U	SMPL
I2159-01A - 1		13843	4.0212	-245.1		PPB	19 Apr 2017 15:04:34	U	SMPL
CCV53 - 1	CCV53	17041	4.9471	-245.1		PPB	19 Apr 2017 15:06:41	U	SMPL
CCB53 - 1	CCB53	-58	-0.0033	-245.1		PPB	19 Apr 2017 15:08:48	U	SMPL