



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86966

Temp.: 21.2 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/19/2017	15:25
2	CAL2	1	Water			7.01	04/19/2017	15:26
3	CAL3	1	Water			10.04	04/19/2017	15:27
4	ICV	1	Water			7.01	04/19/2017	15:28
5	CCV1	1	Water			2.02	04/19/2017	15:30
6	I2755-02	1	Water			6.74	04/19/2017	15:31
7	I2755-02DUP	1	Water			6.75	04/19/2017	15:32
8	I2755-06	1	Water			7.27	04/19/2017	15:33
9	I2758-01	1	Water			11.20	04/19/2017	15:34
10	CCV2	1	Water			12.01	04/19/2017	15:35

2386966

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 12755

WorkList ID : 97575

Date : 4/19/2017 2:57:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/19/2017	Water	I2755-02-4	pH	Cool 4 deg C	PSEG01	P31	MH-41	04/19/2017	9045C
04/19/2017	Water	I2755-06-4	pH	Cool 4 deg C	PSEG01	P31	3652	04/19/2017	9045C
04/19/2017	Water	I2758-01-4	pH	Cool 4 deg C	PSEG01	P31	CLEAN-HARBORS-DRUMS(1	04/19/2017	9045C

Date/Time 04-19-17 2:57 PM
Received by: 
Relinquished by: 

Date/Time 04-19-17 5:23 PM
Received by: 
Relinquished by: JP