



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86968

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	04/19/2017	15:40
2	CAL2	1	Water			7.01	04/19/2017	15:41
3	CAL3	1	Water			10.03	04/19/2017	15:42
4	ICV	1	Water			6.99	04/19/2017	15:43
5	CCV1	1	Water			2.02	04/19/2017	15:45
6	I2744-13	1	Solid	20.03	20	8.77	04/19/2017	15:46
7	I2744-13DUP	1	Solid	20.04	20	8.78	04/19/2017	15:47
8	I2744-14	1	Solid	20.02	20	5.40	04/19/2017	15:48
9	I2744-15	1	Solid	20.01	20	8.30	04/19/2017	15:50
10	I2744-20	1	Solid	20.04	20	6.47	04/19/2017	15:51
11	I2744-21	1	Solid	20.01	20	8.54	04/19/2017	15:52
12	CCV2	1	Water			12.02	04/19/2017	15:53

2886968

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2744 WorkList ID : 97579 Date : 4/19/2017 2:58:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2744-13	Corrosivity	Cool 4 deg C	AKRF02	P62	SB-1(0-2)20170418	04/18/2017	9045C
	Solid	I2744-14	Corrosivity	Cool 4 deg C	AKRF02	P62	SB-3(0-2)20170418	04/18/2017	9045C
	Solid	I2744-15	Corrosivity	Cool 4 deg C	AKRF02	P62	SB-5(0-2)20170418	04/18/2017	9045C
	Solid	I2744-20	Corrosivity	Cool 4 deg C	AKRF02	P62	SB-6(0-2)20170418	04/18/2017	9045C
	Solid	I2744-21	Corrosivity	Cool 4 deg C	AKRF02	P62	SB-4(0-2)20170419	04/19/2017	9045C

Date/Time 04-19-17 3:00PM
 Received by: JP
 Relinquished by: JP

Date/Time 04-19-17 2:30PM
 Received by: JP
 Relinquished by: JP