

LB 86980


INSTRUMENT ID: CV1

7470A

Linear

 μ Abs.:

36554



A= 0.0000e+000

B= 2.7398e-004

C= 2.4824e-002

Rho= 0.9999652

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	0	0.000	0				
0.2	0.200	0.223	0.023	722	0.0 %	722				
2.5	2.500	2.473	-0.027	8936	0.0 %	8936				
5.0	5.000	4.955	-0.045	17996	0.0 %	17996				
7.5	7.500	7.484	-0.016	27226	0.0 %	27226				
10.0	10.000	10.040	0.040	36554	0.0 %	36554				

LB86980 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Apr 2017 10:36:46

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.2	0	-	0.0000	7470A	PPB	20 Apr 2017 10:53:57	S	Std
0.2 - 1	50.2	722	-	0.2000	7470A	PPB	20 Apr 2017 10:56:05	S	Std
2.5 - 1	52.5	8936	-	2.5000	7470A	PPB	20 Apr 2017 10:58:14	S	Std
5.0 - 1	55	17996	-	5.0000	7470A	PPB	20 Apr 2017 11:00:22	S	Std
7.5 - 1	55	27226	-	7.5000	7470A	PPB	20 Apr 2017 11:02:30	S	Std
10.0 - 1	510	36554	-	10.0000	7470A	PPB	20 Apr 2017 11:04:38	S	Std
ICV27 - 1	ICV27	14020	3.8660	-	7470A	PPB	20 Apr 2017 11:07:09	U	SMPL
ICB27 - 1	ICB27	-91	-0.0001	-	7470A	PPB	20 Apr 2017 11:09:15	U	SMPL
CCV54 - 1	CCV54	18209	5.0137	-	7470A	PPB	20 Apr 2017 11:11:21	U	SMPL
CCB54 - 1	CCB54	-89	0.0004	-	7470A	PPB	20 Apr 2017 11:13:28	U	SMPL
CRA - 1	CRA	702	0.2172	-	7470A	PPB	20 Apr 2017 11:15:35	U	SMPL
HighStd - 1	HighStd	35798	9.8327	-	7470A	PPB	20 Apr 2017 11:17:41	U	SMPL
ChkStd - 1	ChkStd	25156	6.9171	-	7470A	PPB	20 Apr 2017 11:19:48	U	SMPL
PB98347BL - 1	PBW	-122	-0.0086	-	7470A	PPB	20 Apr 2017 11:21:56	U	SMPL
PB98347BS - 1	LCSW	13839	3.8164	-	7470A	PPB	20 Apr 2017 11:24:04	U	SMPL
PB98347TB - 1	PB98322TB	-69	0.0059	-	7470A	PPB	20 Apr 2017 11:26:12	U	SMPL
I2751-03 - 1	15-40-COMP	-2	0.0243	-	7470A	PPB	20 Apr 2017 11:28:20	U	SMPL
I2751-06 - 1	RT-1494-COMP	7	0.0267	-	7470A	PPB	20 Apr 2017 11:30:28	U	SMPL
I2751-09 - 1	VNJ-209-COMP	-1	0.0246	-	7470A	PPB	20 Apr 2017 11:32:37	U	SMPL
I2752-03 - 1	15-50-COMP	4	0.0259	-	7470A	PPB	20 Apr 2017 11:34:44	U	SMPL
CCV55 - 1	CCV55	17859	4.9178	-	7470A	PPB	20 Apr 2017 11:36:51	U	SMPL
CCB55 - 1	CCB55	-82	0.0024	-	7470A	PPB	20 Apr 2017 11:38:58	U	SMPL
I2752-06 - 1	VNJ-215-COMP	-6	0.0232	-	7470A	PPB	20 Apr 2017 11:41:05	U	SMPL
I2752-06DUP - 1	VNJ-215-COMPDUP	0	0.0248	-	7470A	PPB	20 Apr 2017 11:43:12	U	SMPL
I2752-06MS - 1	VNJ-215-COMPMS	13291	3.6663	-	7470A	PPB	20 Apr 2017 11:45:19	U	SMPL
I2752-06MSD - 1	VNJ-215-COMPMSD	12878	3.5531	-	7470A	PPB	20 Apr 2017 11:47:26	U	SMPL
PB98358BL - 1	PBW	-51	0.0109	-	7470A	PPB	20 Apr 2017 11:49:33	U	SMPL
PB98358BS - 1	LCSW	13632	3.7597	-	7470A	PPB	20 Apr 2017 11:51:41	U	SMPL
I2755-01 - 1	MH-41	279	0.1013	-	7470A	PPB	20 Apr 2017 11:53:49	U	SMPL
I2755-01DUP - 1	MH-41DUP	222	0.0856	-	7470A	PPB	20 Apr 2017 11:55:57	U	SMPL
I2755-01MS - 1	MH-41MS	15433	4.2532	-	7470A	PPB	20 Apr 2017 11:58:06	U	SMPL
I2755-01MSD - 1	MH-41MSD	15750	4.3400	-	7470A	PPB	20 Apr 2017 12:00:13	U	SMPL
CCV56 - 1	CCV56	17730	4.8825	-	7470A	PPB	20 Apr 2017 12:02:20	U	SMPL
CCB56 - 1	CCB56	-89	0.0004	-	7470A	PPB	20 Apr 2017 12:04:27	U	SMPL
I2755-05 - 1	3652	89	0.0492	-	7470A	PPB	20 Apr 2017 12:06:34	U	SMPL
I2758-01 - 1	CLEAN-HARBORS-DRUMS(1-9)	607	0.1911	-	7470A	PPB	20 Apr 2017 12:08:41	U	SMPL
I2752-06LX5 - 1		-11	0.0218	-	7470A	PPB	20 Apr 2017 12:10:48	U	SMPL
I2752-06A - 1		13587	3.7474	-	7470A	PPB	20 Apr 2017 12:12:55	U	SMPL
I2755-01LX5 - 1		-78	0.0035	-	7470A	PPB	20 Apr 2017 12:15:02	U	SMPL
I2755-01A - 1		14482	3.9926	-	7470A	PPB	20 Apr 2017 12:17:10	U	SMPL
CCV57 - 1	CCV57	17841	4.9129	-	7470A	PPB	20 Apr 2017 12:19:18	U	SMPL
CCB57 - 1	CCB57	-88	0.0007	-	7470A	PPB	20 Apr 2017 12:21:27	U	SMPL