

7471A


LB 86983

INSTRUMENT ID: CV2

Linear

 μ Abs.:

43455



A= 0.0000e+000

B= 2.2811e-004

C= -2.0686e-002

Rho= 0.9996401

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	13	0.000	13				
0.2	0.200	0.192	-0.008	931	0.0 %	931				
2.5	2.500	2.402	-0.098	10623	0.0 %	10623				
5.0	5.000	5.189	0.189	22837	0.0 %	22837				
7.5	7.500	7.542	0.042	33156	0.0 %	33156				
10.0	10.000	9.893	-0.107	43459	0.0 %	43459				

LB86983 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 20 Apr 2017 12:49:10

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	13	-	0.00007471A	PPB		20 Apr 2017 12:50:39	S	Std
0.2 - 1	50.2	931	-	0.20007471A	PPB		20 Apr 2017 12:52:57	S	Std
2.5 - 1	52.5	10623	-	2.50007471A	PPB		20 Apr 2017 12:55:14	S	Std
5.0 - 1	55	22837	-	5.00007471A	PPB		20 Apr 2017 12:57:32	S	Std
7.5 - 1	57.5	33156	-	7.50007471A	PPB		20 Apr 2017 12:59:49	S	Std
10.0 - 1	510	43459	-	10.00007471A	PPB		20 Apr 2017 13:04:46	S	Std
ICV28 - 1	ICV28	17409	3.9504	-7471A	PPB		20 Apr 2017 13:08:03	U	SMPL
ICB28 - 1	ICB28	-58	-0.0339	-7471A	PPB		20 Apr 2017 13:10:19	U	SMPL
CCV58 - 1	CCV58	22347	5.0768	-7471A	PPB		20 Apr 2017 13:12:34	U	SMPL
CCB58 - 1	CCB58	-90	-0.0412	-7471A	PPB		20 Apr 2017 13:14:50	U	SMPL
CRA - 1	CRA	905	0.1858	-7471A	PPB		20 Apr 2017 13:17:06	U	SMPL
HighStd - 1	HighStd	44305	10.0856	-7471A	PPB		20 Apr 2017 13:19:22	U	SMPL
ChkStd - 1	ChkStd	30524	6.9420	-7471A	PPB		20 Apr 2017 13:21:38	U	SMPL
PB98348BL - 1	PBS	-98	-0.0430	-7471A	PPB		20 Apr 2017 13:23:55	U	SMPL
PB98348BS - 1	LCSS	14885	3.3747	-7471A	PPB		20 Apr 2017 13:26:12	U	SMPL
I2749-01 - 1	E2-V6-WSW	78409	17.8649	-7471A	PPB		20 Apr 2017 13:28:29	U	SMPL
I2749-02 - 1	E2-V6-NSW-01	3786	0.8429	-7471A	PPB		20 Apr 2017 13:30:46	U	SMPL
I2749-03 - 1	E2-V6-NSW-02	101041	23.0274	-7471A	PPB		20 Apr 2017 13:33:05	U	SMPL
I2750-01 - 1	E2-U5-SSW	21687	4.9263	-7471A	PPB		20 Apr 2017 13:35:23	U	SMPL
I2750-02 - 1	E2-U5-BASE	18730	4.2517	-7471A	PPB		20 Apr 2017 13:37:40	U	SMPL
CCV59 - 1	CCV59	22365	5.0809	-7471A	PPB		20 Apr 2017 13:39:56	U	SMPL
CCB59 - 1	CCB59	-11	-0.0232	-7471A	PPB		20 Apr 2017 13:42:12	U	SMPL
I2750-02DUP - 1	E2-U5-BASEDUP	16042	3.6386	-7471A	PPB		20 Apr 2017 13:44:27	U	SMPL
I2750-02MS - 1	E2-U5-BASEMS	35731	8.1298	-7471A	PPB		20 Apr 2017 13:49:32	U	SMPL
I2750-02MSD - 1	E2-U5-BASEMSD	34307	7.8050	-7471A	PPB		20 Apr 2017 13:51:47	U	SMPL
I2750-03 - 1	E2-U6-BASE	123246	28.0925	-7471A	PPB		20 Apr 2017 13:54:03	U	SMPL
I2750-04 - 1	E2-T5-WSW	708165	HIGH	-7471A	PPB		20 Apr 2017 13:56:20	X	
I2750-05 - 1	E2-T5-BASE	596	0.1153	-7471A	PPB		20 Apr 2017 13:58:12	U	SMPL
I2750-06 - 1	E2-T5-NSW	6232	1.4009	-7471A	PPB		20 Apr 2017 14:00:33	U	SMPL
I2751-01 - 1	15-40-COMP	3015	0.6671	-7471A	PPB		20 Apr 2017 14:02:50	U	SMPL
I2751-04 - 1	RT-1494-COMP	2763	0.6096	-7471A	PPB		20 Apr 2017 14:05:08	U	SMPL
I2751-07 - 1	VNJ-209-COMP	12651	2.8651	-7471A	PPB		20 Apr 2017 14:07:25	U	SMPL
CCV60 - 1	CCV60	22350	5.0775	-7471A	PPB		20 Apr 2017 14:09:41	U	SMPL
CCB60 - 1	CCB60	-19	-0.0250	-7471A	PPB		20 Apr 2017 14:11:57	U	SMPL
I2752-01 - 1	15-50-COMP	1703	0.3678	-7471A	PPB		20 Apr 2017 14:14:13	U	SMPL
I2752-04 - 1	VNJ-215-COMP	2998	0.6632	-7471A	PPB		20 Apr 2017 14:16:30	U	SMPL
PB98357BL - 1	PBS	-3	-0.0214	-7471A	PPB		20 Apr 2017 14:18:46	U	SMPL
PB98357BS - 1	LCSS	16067	3.6443	-7471A	PPB		20 Apr 2017 14:21:03	U	SMPL
I2744-01 - 1	SB-2(8-10)20170418	750	0.1504	-7471A	PPB		20 Apr 2017 14:23:20	U	SMPL
I2744-02 - 1	SB-2(18-20)20170418	371	0.0639	-7471A	PPB		20 Apr 2017 14:25:36	U	SMPL
I2744-03 - 1	SB-9(11-13)20170418	266	0.0400	-7471A	PPB		20 Apr 2017 14:27:53	U	SMPL
I2744-04 - 1	SB-9(18-20)20170418	266	0.0400	-7471A	PPB		20 Apr 2017 14:30:11	U	SMPL
I2744-05 - 1	SB-1(0-2)20170418	5577	1.2515	-7471A	PPB		20 Apr 2017 14:32:29	U	SMPL
I2744-06 - 1	SB-1(5-7)20170418	583	0.1123	-7471A	PPB		20 Apr 2017 14:34:45	U	SMPL
CCV61 - 1	CCV61	22407	5.0905	-7471A	PPB		20 Apr 2017 14:37:01	U	SMPL
CCB61 - 1	CCB61	-27	-0.0268	-7471A	PPB		20 Apr 2017 14:39:18	U	SMPL
I2744-07 - 1	SB-7(5-7)20170418	483	0.0895	-7471A	PPB		20 Apr 2017 14:41:34	U	SMPL
I2744-08 - 1	SB-7(18-20)20170418	185	0.0215	-7471A	PPB		20 Apr 2017 14:43:50	U	SMPL
I2744-09 - 1	SB-3(0-2)20170418	3965	0.8838	-7471A	PPB		20 Apr 2017 14:46:07	U	SMPL
I2744-10 - 1	SB-3(13-15)20170418	377	0.0653	-7471A	PPB		20 Apr 2017 14:48:23	U	SMPL
I2744-11 - 1	SB-5(0-2)20170418	15845	3.5937	-7471A	PPB		20 Apr 2017 14:50:40	U	SMPL
I2744-12 - 1	SB-5(13-15)20170418	251	0.0366	-7471A	PPB		20 Apr 2017 14:52:57	U	SMPL
I2744-16 - 1	SB-6(0-2)20170418	3147	0.6972	-7471A	PPB		20 Apr 2017 14:55:14	U	SMPL
I2744-17 - 1	SB-6(13-15)20170418	314	0.0509	-7471A	PPB		20 Apr 2017 14:57:32	U	SMPL
I2744-18 - 1	SB-4(0-2)20170419	35092	7.9840	-7471A	PPB		20 Apr 2017 14:59:49	U	SMPL
I2744-18DUP - 1	SB-4(0-2)20170419DUP	32671	7.4318	-7471A	PPB		20 Apr 2017 15:04:46	U	SMPL
CCV62 - 1	CCV62	22884	5.1993	-7471A	PPB		20 Apr 2017 15:07:02	U	SMPL
CCB62 - 1	CCB62	-13	-0.0237	-7471A	PPB		20 Apr 2017 15:09:19	U	SMPL
I2744-18MS - 1	SB-4(0-2)20170419MS	49913	11.3648	-7471A	PPB		20 Apr 2017 15:11:35	U	SMPL
I2744-18MSD - 1	SB-4(0-2)20170419MSD	50003	11.3853	-7471A	PPB		20 Apr 2017 15:16:40	U	SMPL
I2744-19 - 1	SB-4(13-15)20170419	639	0.1251	-7471A	PPB		20 Apr 2017 15:18:56	U	SMPL
I2750-02LX5 - 1		3560	0.7914	-7471A	PPB		20 Apr 2017 15:21:12	U	SMPL
I2750-02A - 1		36290	8.2573	-7471A	PPB		20 Apr 2017 15:23:28	U	SMPL
I2744-18LX5 - 1		6729	1.5142	-7471A	PPB		20 Apr 2017 15:25:45	U	SMPL
I2744-18A - 1		52107	11.8652	-7471A	PPB		20 Apr 2017 15:28:02	U	SMPL
I2750-04DLX10 - 1		128760	29.3503	-7471A	PPB		20 Apr 2017 15:30:19	U	SMPL
CCV63 - 1	CCV63	22795	5.1790	-7471A	PPB		20 Apr 2017 15:32:37	U	SMPL
CCB63 - 1	CCB63	-26	-0.0266	-7471A	PPB		20 Apr 2017 15:34:55	U	SMPL
I2749-01DLX5 - 1		15726	3.5665	-7471A	PPB		20 Apr 2017 15:37:12	U	SMPL
I2749-03DLX5 - 1		20354	4.6222	-7471A	PPB		20 Apr 2017 15:39:28	U	SMPL
I2750-03DLX5 - 1		25048	5.6929	-7471A	PPB		20 Apr 2017 15:41:45	U	SMPL
-2750-04DLX2X50 - 1		25885	5.8838	-7471A	PPB		20 Apr 2017 15:44:02	U	SMPL
I2744-18DLX2 - 1		17685	4.0134	-7471A	PPB		20 Apr 2017 15:46:18	U	SMPL
I2744-18DUPDLX2 - 1		17305	3.9267	-7471A	PPB		20 Apr 2017 15:51:20	U	SMPL
I2744-18MSDLX2 - 1		26592	6.0451	-7471A	PPB		20 Apr 2017 15:53:36	U	SMPL
I2744-18MSDDLX2 - 1		26733	6.0773	-7471A	PPB		20 Apr 2017 15:55:53	U	SMPL
I2744-18LDLX10 - 1		3619	0.8048	-7471A	PPB		20 Apr 2017 15:58:10	U	SMPL
I2744-18ADLX2 - 1		25973	5.9039	-7471A	PPB		20 Apr 2017 16:00:28	U	SMPL
CCV64 - 1	CCV64	23114	5.2518	-7471A	PPB		20 Apr 2017 16:02:46	U	SMPL
CCB64 - 1	CCB64	5	-0.0195	-7471A	PPB		20 Apr 2017 16:05:03	U	SMPL