



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB86994

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	04/20/2017	14:30
2	I2755-03		1	>100.0	>212.0	04/20/2017	15:00
3	I2755-03DUP		1	>100.0	>212.0	04/20/2017	15:15
4	I2755-07		1	>100.0	>212.0	04/20/2017	15:30
5	I2758-02		1	>100.0	>212.0	04/20/2017	15:50

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 86994

Worklist Name : FP42017

Worklist ID : 97619

Date : 4/20/2017 2:10:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/21/2017	Water	12755-03	BC Flash Point	Cool 4 deg C	PSEG01	P31	MH-41	04/19/2017	1010A
04/21/2017	Water	12755-07	BC Flash Point	Cool 4 deg C	PSEG01	P31	3652	04/19/2017	1010A
04/26/2017	Water	12758-02	EA Flash Point	Cool 4 deg C	PSEG01	P31	CLEAN-HARBORS-DRUMS(1	04/19/2017	1010A

KS

Date/Time

4/20/17 2:15 PM

Received by:

KS

Relinquished by:

CP

Date/Time

4/20/17 16:00

Received by:

CP

Relinquished by:

KS