


L1386995

ISM02.4

INSTRUMENT ID: CV1

μ Abs.:

44752



Linear

A= 0.0000e+000

B= 2.2342e-004

C= -6.2088e-003

Rho= 0.9999738

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.001	0.001	31	0.000	31					1
0.2	0.200	0.207	0.007	954	0.0 %	954					3
2.5	2.500	2.511	0.011	11266	0.0 %	11266					0
5.0	5.000	4.950	-0.050	22183	0.0 %	22183					-1
7.5	7.500	7.539	0.039	33773	0.0 %	33773					1
10.0	10.000	9.992	-0.008	44752	0.0 %	44752					0

ns3
4/20/17

LB86995

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	31	-		0 ISM02.4	PPB	4/20/2017 16:29	Std
	0.2 S0.2	954	-		0.2 ISM02.4	PPB	4/20/2017 16:31	Std
	2.5 S2.5	11266	-		2.5 ISM02.4	PPB	4/20/2017 16:34	Std
	5 S5	22183	-		5 ISM02.4	PPB	4/20/2017 16:36	Std
	7.5 S7.5	33773	-		7.5 ISM02.4	PPB	4/20/2017 16:38	Std
	10 S10	44752	-		10 ISM02.4	PPB	4/20/2017 16:41	Std
ICV	ICV	16333	3.6429 -		ISM02.4	PPB	4/20/2017 16:43	SMPL
ICB	ICB	-73	-0.0225 -		ISM02.4	PPB	4/20/2017 16:45	SMPL
CCV065	CCV065	21904	4.8876 -		ISM02.4	PPB	4/20/2017 16:48	SMPL
CCB065	CCB065	-75	-0.023 -		ISM02.4	PPB	4/20/2017 16:50	SMPL
PB98349BL	PBW003	9	-0.0042 -		ISM02.4	PPB	4/20/2017 16:52	SMPL
I2689-03	MF4A03	-3	-0.0069 -		ISM02.4	PPB	4/20/2017 16:55	SMPL
I2689-04	MF4A04	19	-0.002 -		ISM02.4	PPB	4/20/2017 16:57	SMPL
CCV066	CCV066	21792	4.8626 -		ISM02.4	PPB	4/20/2017 16:59	SMPL
CCB066	CCB066	-36	-0.0143 -		ISM02.4	PPB	4/20/2017 17:01	SMPL