



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87002

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 99.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/21/2017	11:50
2	CAL2	1	Water			7.00	04/21/2017	11:51
3	CAL3	1	Water			10.02	04/21/2017	11:52
4	ICV	1	Water			7.00	04/21/2017	11:53
5	CCV1	1	Water			2.02	04/21/2017	12:10
6	I2790-03	1	Solid	20.3	20	8.53	04/21/2017	12:11
7	I2790-03DUP	1	Solid	20.4	20	8.54	04/21/2017	12:12
8	I2790-06	1	Solid	20.2	20	8.67	04/21/2017	12:13
9	CCV2	1	Water			12.02	04/21/2017	12:15

1387002

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2790 WorkList ID : 97646 Date : 4/21/2017 11:38:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/26/2017	Solid	I2790-03 -A	Corrosivity	Cool 4 deg C	PSEG01	P51	ROLLOFFS-156-VNJ213-155-	04/21/2017	9045C
04/26/2017	Solid	I2790-06 -A	Corrosivity	Cool 4 deg C	PSEG01	P51	ROLLOFFS-VNJ238-VNJ205-	04/21/2017	9045C

Date/Time 04-21-17 11:40 AM
Received by: 78
Relinquished by: [Signature]

Date/Time 04-21-17 530 PM
Received by: [Signature]
Relinquished by: JP