



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87004

Temp.: 18.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/21/2017	11:50
2	CAL2	1	Water			7.00	04/21/2017	11:51
3	CAL3	1	Water			10.02	04/21/2017	11:52
4	ICV	1	Water			7.00	04/21/2017	11:53
5	CCV1	1	Water			2.01	04/21/2017	11:54
6	I2792-01	1	Water			7.43	04/21/2017	11:55
7	I2792-01DUP	1	Water			7.44	04/21/2017	11:56
8	CCV2	1	Water			12.01	04/21/2017	11:57

LB 87004

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i2792

WorkList ID : 97650

Date : 4/21/2017 11:39:15 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/21/2017	Water	I2792-01	pH	Cool 4 deg C	PSEG01	P51	RINSE-WATER	04/21/2017	9045C

Date/Time 04-21-17 11:39 AM
Received by: JI
Relinquished by: JD

Date/Time 04-21-17 11:30 AM
Received by: JD
Relinquished by: JD