


7470A

LB87005
INSTRUMENT ID: CV2

Linear

 μ Abs.:

34710



A= 0.0000e+000

B= 2.8988e-004

C= -3.4303e-002

Rho= 0.9998416

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.021	0.021	191	0.000	191				
0.2	0.200	0.221	0.021	880	0.0 %	880				
2.5	2.500	2.530	0.030	8845	0.0 %	8845				
5.0	5.000	4.854	-0.146	16865	0.0 %	16865				
7.5	7.500	7.547	0.047	26153	0.0 %	26153				
10.0	10.000	10.027	0.027	34710	0.0 %	34710				

LB87005 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Apr 2017 12:04:43

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	191	-	0.0000	7470A	PPB	21 Apr 2017 12:17:43	S	Std
0.2 - 1	50.2	880	-	0.2000	7470A	PPB	21 Apr 2017 12:19:51	S	Std
2.5 - 1	52.5	8845	-	2.5000	7470A	PPB	21 Apr 2017 12:21:59	S	Std
5.0 - 1	55	16865	-	5.0000	7470A	PPB	21 Apr 2017 12:24:08	S	Std
7.5 - 1	57.5	26153	-	7.5000	7470A	PPB	21 Apr 2017 12:26:16	S	Std
10.0 - 1	510	34710	-	10.0000	7470A	PPB	21 Apr 2017 12:30:46	S	Std
ICV29 - 1	ICV29	13585	3.9037	-	7470A	PPB	21 Apr 2017 12:33:16	U	SMPL
ICB29 - 1	ICB29	143	0.0071	-	7470A	PPB	21 Apr 2017 12:35:22	U	SMPL
CCV67 - 1	CCV67	17656	5.0837	-	7470A	PPB	21 Apr 2017 12:37:29	U	SMPL
CCB67 - 1	CCB67	138	0.0057	-	7470A	PPB	21 Apr 2017 12:39:35	U	SMPL
CRA - 1	CRA	851	0.2124	-	7470A	PPB	21 Apr 2017 12:41:42	U	SMPL
HighStd - 1	HighStd	34874	10.0748	-	7470A	PPB	21 Apr 2017 12:43:48	U	SMPL
ChkStd - 1	ChkStd	25283	7.2946	-	7470A	PPB	21 Apr 2017 12:45:55	U	SMPL
PB98375BL - 1	PBW	25	-0.0271	-	7470A	PPB	21 Apr 2017 12:48:04	U	SMPL
PB98375BS - 1	LCSW	13590	3.9051	-	7470A	PPB	21 Apr 2017 12:50:12	U	SMPL
PB98375TB - 1	PB98351TB	71	-0.0137	-	7470A	PPB	21 Apr 2017 12:52:19	U	SMPL
I2755-03 - 1	MH-41	108	-0.0030	-	7470A	PPB	21 Apr 2017 12:54:28	U	SMPL
I2755-07 - 1	3652	156	0.0109	-	7470A	PPB	21 Apr 2017 12:56:36	U	SMPL
I2758-02 - 1	CLEAN-HARBORS-DRUMS(1-9)	157	0.0112	-	7470A	PPB	21 Apr 2017 12:58:45	U	SMPL
I2758-02DUP - 1	CLEAN-HARBORS-DRUMS(1-9)DUP	228	0.0318	-	7470A	PPB	21 Apr 2017 13:00:52	U	SMPL
CCV68 - 1	CCV68	17473	5.0307	-	7470A	PPB	21 Apr 2017 13:02:59	U	SMPL
CCB68 - 1	CCB68	143	0.0071	-	7470A	PPB	21 Apr 2017 13:05:05	U	SMPL
I2758-02MS - 1	CLEAN-HARBORS-DRUMS(1-9)MS	13450	3.8645	-	7470A	PPB	21 Apr 2017 13:07:12	U	SMPL
I2758-02MSD - 1	CLEAN-HARBORS-DRUMS(1-9)MSD	13277	3.8144	-	7470A	PPB	21 Apr 2017 13:09:19	U	SMPL
I2758-02LX5 - 1		86	-0.0094	-	7470A	PPB	21 Apr 2017 13:11:25	U	SMPL
I2758-02A - 1		13679	3.9309	-	7470A	PPB	21 Apr 2017 13:13:32	U	SMPL
CCV69 - 1	CCV69	17640	5.0791	-	7470A	PPB	21 Apr 2017 13:15:40	U	SMPL
CCB69 - 1	CCB69	145	0.0077	-	7470A	PPB	21 Apr 2017 13:17:48	U	SMPL