



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87006

Temp.: 21.2 °C

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/21/2017	12:30
2	CAL2	1	Water			7.01	04/21/2017	12:31
3	CAL3	1	Water			10.03	04/21/2017	12:32
4	ICV	1	Water			7.00	04/21/2017	12:33
5	CCV1	1	Water			2.02	04/21/2017	12:34
6	I2781-01	1	Water			5.60	04/21/2017	12:35
7	I2781-01DUP	1	Water			5.61	04/21/2017	12:36
8	I2781-02	1	Water			5.50	04/21/2017	12:37
9	CCV2	1	Water			12.02	04/21/2017	12:39

1387006

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i2781 WorkList ID : 97652 Date : 4/21/2017 12:23:59 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2781-01 -A	pH	Cool 4 deg C	STER07	P11	CELL-A	04/19/2017	SM 4500-PH B
	Water	I2781-02 -A	pH	Cool 4 deg C	STER07	P11	CELLS-B-C	04/19/2017	SM 4500-PH B

Date/Time 04/21/17 12:20 PM
Received by: JP
Relinquished by: JP

Date/Time 04/21/17 21:00 PM
Received by: CP
Relinquished by: JP