


7471A

LB87011
INSTRUMENT ID: CV1

Linear

 μ Abs.:

46532



A= 0.0000e+000

B= 2.1588e-004

C= 2.2215e-002

Rho= 0.9999253

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.039	0.039	78	0.000	78				
0.2	0.200	0.217	0.017	902	0.0 %	902				
2.5	2.500	2.466	-0.034	11319	0.0 %	11319				
5.0	5.000	4.974	-0.026	22939	0.0 %	22939				
7.5	7.500	7.437	-0.063	34346	0.0 %	34346				
10.0	10.000	10.067	0.067	46532	0.0 %	46532				

LB87011 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 21 Apr 2017 15:33:29

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.0	78	-	0.00007471A	PPB		21 Apr 2017 15:36:33	S	Std
0.2 - 1	50.2	902	-	0.20007471A	PPB		21 Apr 2017 15:38:51	S	Std
2.5 - 1	52.5	11319	-	2.50007471A	PPB		21 Apr 2017 15:41:08	S	Std
5.0 - 1	55	22939	-	5.00007471A	PPB		21 Apr 2017 15:43:26	S	Std
7.5 - 1	57.5	34346	-	7.50007471A	PPB		21 Apr 2017 15:45:43	S	Std
10.0 - 1	510	46532	-	10.00007471A	PPB		21 Apr 2017 15:50:35	S	Std
ICV30 - 1	ICV30	17792	3.8631	-7471A	PPB		21 Apr 2017 15:53:48	U	SMPL
ICB30 - 1	ICB30	-30	0.0157	-7471A	PPB		21 Apr 2017 15:56:03	U	SMPL
CCV70 - 1	CCV70	22797	4.9436	-7471A	PPB		21 Apr 2017 15:58:19	U	SMPL
CCB70 - 1	CCB70	-38	0.0140	-7471A	PPB		21 Apr 2017 16:00:35	U	SMPL
CRA - 1	CRA	867	0.2094	-7471A	PPB		21 Apr 2017 16:02:50	U	SMPL
HighStd - 1	HighStd	47014	10.1714	-7471A	PPB		21 Apr 2017 16:05:06	U	SMPL
ChkStd - 1	ChkStd	32017	6.9339	-7471A	PPB		21 Apr 2017 16:07:22	U	SMPL
PB98398BL - 1	PBS	-64	0.0084	-7471A	PPB		21 Apr 2017 16:09:40	U	SMPL
PB98398BS - 1	LCSS	17057	3.7044	-7471A	PPB		21 Apr 2017 16:14:30	U	SMPL
I2774-01 - 1	WETLANDSAND-PPC	418	0.1125	-7471A	PPB		21 Apr 2017 16:21:37	U	SMPL
I2774-01DUP - 1	WETLANDSAND-PPCDUP	352	0.0982	-7471A	PPB		21 Apr 2017 16:23:53	U	SMPL
I2774-01MS - 1	WETLANDSAND-PPCMS	16201	3.5196	-7471A	PPB		21 Apr 2017 16:26:11	U	SMPL
I2774-01MSD - 1	WETLANDSAND-PPCMSD	16266	3.5337	-7471A	PPB		21 Apr 2017 16:28:29	U	SMPL
I2783-01 - 1	C-9	624	0.1569	-7471A	PPB		21 Apr 2017 16:30:45	U	SMPL
CCV71 - 1	CCV71	22531	4.8861	-7471A	PPB		21 Apr 2017 16:33:01	U	SMPL
CCB71 - 1	CCB71	-11	0.0198	-7471A	PPB		21 Apr 2017 16:35:17	U	SMPL
I2783-03 - 1	C-10	1170	0.2748	-7471A	PPB		21 Apr 2017 16:37:33	U	SMPL
I2783-05 - 1	C-11	966	0.2308	-7471A	PPB		21 Apr 2017 16:39:49	U	SMPL
I2784-01 - 1	FK-GF-02-005-A	3448	0.7666	-7471A	PPB		21 Apr 2017 16:42:05	U	SMPL
I2784-03 - 1	FK-GF-02-005-B	3903	0.8648	-7471A	PPB		21 Apr 2017 16:44:21	U	SMPL
I2790-01 - 1	ROLLOFFS-156-VNJ213-155-160-342-COMP	9568	2.0877	-7471A	PPB		21 Apr 2017 16:46:38	U	SMPL
I2790-04 - 1	ROLLOFFS-VNJ238-VNJ205-VNJ221-VNJ249-COMP	10416	2.2708	-7471A	PPB		21 Apr 2017 16:48:55	U	SMPL
I2808-01 - 1	IM-TS-EME-09	2036	0.4617	-7471A	PPB		21 Apr 2017 16:51:13	U	SMPL
I2774-01LX5 - 1		87	0.0410	-7471A	PPB		21 Apr 2017 16:53:30	U	SMPL
I2774-01A - 1		18429	4.0006	-7471A	PPB		21 Apr 2017 16:55:48	U	SMPL
CCV72 - 1	CCV72	22950	4.9766	-7471A	PPB		21 Apr 2017 16:58:04	U	SMPL
CCB72 - 1	CCB72	10	0.0244	-7471A	PPB		21 Apr 2017 17:00:20	U	SMPL