


7470A

LB 87024
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35325



A= 0.0000e+000

B= 2.8208e-004

C= -2.0051e-002

Rho= 0.9998206

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	137	0.000	137				
0.2	0.200	0.241	0.041	926	0.0 %	926				
2.5	2.500	2.403	-0.097	8589	0.0 %	8589				
5.0	5.000	4.975	-0.025	17707	0.0 %	17707				
7.5	7.500	7.619	0.119	27080	0.0 %	27080				
10.0	10.000	9.944	-0.056	35325	0.0 %	35325				

LB87024 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Apr 2017 14:46:01

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	137	-	0.0007470A	PPB		24 Apr 2017 14:53:03	S	Std
0.2 - 1	80.2	926	-	0.20007470A	PPB		24 Apr 2017 14:55:11	S	Std
2.5 - 1	82.5	8589	-	2.50007470A	PPB		24 Apr 2017 14:57:19	S	Std
5.0 - 1	85	17707	-	5.00007470A	PPB		24 Apr 2017 14:59:27	S	Std
7.5 - 1	87.5	27080	-	7.50007470A	PPB		24 Apr 2017 15:01:35	S	Std
10.0 - 1	87.5	35325	-	10.00007470A	PPB		24 Apr 2017 15:03:42	S	Std
ICV31 - 1	ICV31	13727	3.8520	-7470A	PPB		24 Apr 2017 15:06:26	U	SMPL
ICB31 - 1	ICB31	65	-0.0017	-7470A	PPB		24 Apr 2017 15:08:33	U	SMPL
CCV73 - 1	CCV73	18381	5.1648	-7470A	PPB		24 Apr 2017 15:10:39	U	SMPL
CCB73 - 1	CCB73	73	0.0005	-7470A	PPB		24 Apr 2017 15:12:46	U	SMPL
CRA - 1	CRA	896	0.2327	-7470A	PPB		24 Apr 2017 15:14:52	U	SMPL
HighStd - 1	HighStd	35396	9.9643	-7470A	PPB		24 Apr 2017 15:16:59	U	SMPL
ChkStd - 1	ChkStd	24925	7.0107	-7470A	PPB		24 Apr 2017 15:19:06	U	SMPL
PB98427BL - 1	PBW	-25	-0.0271	-7470A	PPB		24 Apr 2017 15:21:13	U	SMPL
PB98427BS - 1	LCSW	13910	3.9036	-7470A	PPB		24 Apr 2017 15:23:21	U	SMPL
PB98427TB - 1	PB98382TB	15	-0.0158	-7470A	PPB		24 Apr 2017 15:25:29	U	SMPL
I2780-02 - 1	PE-1	1980	0.5385	-7470A	PPB		24 Apr 2017 15:27:37	U	SMPL
I2780-04 - 1	PE-2	674	0.1701	-7470A	PPB		24 Apr 2017 15:29:45	U	SMPL
I2780-06 - 1	PE-3	386	0.0888	-7470A	PPB		24 Apr 2017 15:31:53	U	SMPL
I2780-08 - 1	PE-4	341	0.0761	-7470A	PPB		24 Apr 2017 15:34:00	U	SMPL
CCV74 - 1	CCV74	18632	5.2356	-7470A	PPB		24 Apr 2017 15:36:06	U	SMPL
CCB74 - 1	CCB74	131	0.0169	-7470A	PPB		24 Apr 2017 15:38:13	U	SMPL
I2780-08DUP - 1	PE-4DUP	383	0.0880	-7470A	PPB		24 Apr 2017 15:42:51	U	SMPL
I2780-08MS - 1	PE-4MS	13520	3.7936	-7470A	PPB		24 Apr 2017 15:44:57	U	SMPL
I2780-08MSD - 1	PE-4MSD	13563	3.8058	-7470A	PPB		24 Apr 2017 15:47:04	U	SMPL
I2780-10 - 1	PE-5	442	0.1046	-7470A	PPB		24 Apr 2017 15:49:11	U	SMPL
I2780-11 - 1	B1A	548	0.1345	-7470A	PPB		24 Apr 2017 15:51:18	U	SMPL
I2780-12 - 1	B1B	561	0.1382	-7470A	PPB		24 Apr 2017 15:53:26	U	SMPL
PB98428BL - 1	PBW	118	0.0132	-7470A	PPB		24 Apr 2017 15:55:34	U	SMPL
PB98428BS - 1	LCSW	14060	3.9459	-7470A	PPB		24 Apr 2017 15:57:43	U	SMPL
I2792-01 - 1	RINSE-WATER	817	0.2104	-7470A	PPB		24 Apr 2017 15:59:51	U	SMPL
I2802-02 - 1	MH-6	200	0.0364	-7470A	PPB		24 Apr 2017 16:01:58	U	SMPL
CCV75 - 1	CCV75	18620	5.2322	-7470A	PPB		24 Apr 2017 16:04:06	U	SMPL
CCB75 - 1	CCB75	129	0.0163	-7470A	PPB		24 Apr 2017 16:06:13	U	SMPL
I2802-02DUP - 1	MH-6DUP	305	0.0660	-7470A	PPB		24 Apr 2017 16:08:20	U	SMPL
I2802-02MS - 1	MH-6MS	14215	3.9897	-7470A	PPB		24 Apr 2017 16:10:26	U	SMPL
I2802-02MSD - 1	MH-6MSD	14004	3.9301	-7470A	PPB		24 Apr 2017 16:12:33	U	SMPL
I2812-02 - 1	WATER	105	0.0096	-7470A	PPB		24 Apr 2017 16:14:41	U	SMPL
PB98429BL - 1	PBW	164	0.0262	-7470A	PPB		24 Apr 2017 16:16:48	U	SMPL
PB98429BS - 1	LCSW	15485	4.3479	-7470A	PPB		24 Apr 2017 16:18:56	U	SMPL
PB98429TB - 1	PB98400TB	105	0.0096	-7470A	PPB		24 Apr 2017 16:21:04	U	SMPL
I2790-03 - 1	ROLLOFFS-156-VNJ213-155-160-342-COMP	142	0.0200	-7470A	PPB		24 Apr 2017 16:23:12	U	SMPL
I2790-06 - 1	ROLLOFFS-VNJ238-VNJ205-VNJ221-VNJ249-COMP	131	0.0169	-7470A	PPB		24 Apr 2017 16:25:21	U	SMPL
I2790-06DUP - 1	ROLLOFFS-VNJ238-VNJ205-VNJ221-VNJ249-COMP	134	0.0177	-7470A	PPB		24 Apr 2017 16:27:28	U	SMPL
CCV76 - 1	CCV76	18492	5.1961	-7470A	PPB		24 Apr 2017 16:29:35	U	SMPL
CCB76 - 1	CCB76	55	-0.0045	-7470A	PPB		24 Apr 2017 16:31:42	U	SMPL
I2790-06MS - 1	ROLLOFFS-VNJ238-VNJ205-VNJ221-VNJ249-COMPMS	13566	3.8066	-7470A	PPB		24 Apr 2017 16:33:49	U	SMPL
I2790-06MSD - 1	ROLLOFFS-VNJ238-VNJ205-VNJ221-VNJ249-COMPMSD	13216	3.7079	-7470A	PPB		24 Apr 2017 16:35:57	U	SMPL
I2792-02 - 1	RINSE-WATER	118	0.0132	-7470A	PPB		24 Apr 2017 16:38:04	U	SMPL
I2812-01 - 1	SOIL	176	0.0296	-7470A	PPB		24 Apr 2017 16:40:11	U	SMPL
I2780-08LX5 - 1		316	0.0691	-7470A	PPB		24 Apr 2017 16:42:19	U	SMPL
I2780-08A - 1		14310	4.0165	-7470A	PPB		24 Apr 2017 16:44:26	U	SMPL
I2802-02LX5 - 1		213	0.0400	-7470A	PPB		24 Apr 2017 16:46:34	U	SMPL
I2802-02A - 1		14561	4.0873	-7470A	PPB		24 Apr 2017 16:48:43	U	SMPL
I2790-06LX5 - 1		104	0.0093	-7470A	PPB		24 Apr 2017 16:50:52	U	SMPL
I2790-06A - 1		14310	4.0165	-7470A	PPB		24 Apr 2017 16:52:59	U	SMPL
CCV77 - 1	CCV77	17476	4.9095	-7470A	PPB		24 Apr 2017 16:55:06	U	SMPL
CCB77 - 1	CCB77	51	-0.0057	-7470A	PPB		24 Apr 2017 16:57:13	U	SMPL