



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87035

Temp.: 23.2 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/24/2017	16:30
2	CAL2	1	Water			7.01	04/24/2017	16:31
3	CAL3	1	Water			10.03	04/24/2017	16:32
4	ICV	1	Water			7.00	04/24/2017	16:33
5	CCV1	1	Water			2.01	04/24/2017	16:35
6	I2832-05	1	Solid	20.02	20	7.91	04/24/2017	16:36
7	I2832-05DUP	1	Solid	20.04	20	7.92	04/24/2017	16:37
8	I2832-11	1	Solid	20.02	20	7.89	04/24/2017	16:38
9	CCV2	1	Water			12.01	04/24/2017	16:39

2387035

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2832 WorkList ID : 97703 Date : 4/24/2017 4:19:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/27/2017	Solid	I2832-05	Corrosivity	Cool 4 deg C	PSEG01	P52	TEST-PIT-RD-(0-3)	04/24/2017	9045C
04/27/2017	Solid	I2832-11	Corrosivity	Cool 4 deg C	PSEG01	P52	TEST-PIT-CURVE-(0-5)	04/24/2017	9045C

Date/Time 04-24-17 4:19 PM
Received by: JP
Relinquished by: JP

Date/Time 04-24-17 5:30 PM
Received by: JP
Relinquished by: JP