



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87036

Temp.: 23.2 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/24/2017	16:30
2	CAL2	1	Water			7.01	04/24/2017	16:31
3	CAL3	1	Water			10.03	04/24/2017	16:32
4	ICV	1	Water			7.00	04/24/2017	16:33
5	CCV1	1	Water			2.02	04/24/2017	16:50
6	I2825-01	1	Solid	20.03	20	7.18	04/24/2017	16:51
7	I2825-01DUP	1	Solid	20.04	20	7.19	04/24/2017	16:52
8	I2828-01	1	Solid	20.03	20	7.19	04/24/2017	16:53
9	CCV2	1	Water			12.02	04/24/2017	16:55

WORKLIST(Hardcopy Internal Chain)

LD 87036

WorkList Name : ph i2825

WorkList ID : 97706

Date : 4/24/2017 4:45:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2825-01	pH	Cool 4 deg C	TULL02	P41	EME-COMPOST-2	04/24/2017	9045C
	Solid	I2828-01	pH	Cool 4 deg C	TULL02	P42	EME-COMPOST-02	04/24/2017	9045C

Date/Time 04-24-17 11:45 AM
 Received by: SR
 Relinquished by: SR

Date/Time 04-24-17 G-10013
 Received by: SR
 Relinquished by: SR